

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101018\
 Data File : PO049850.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Oct 2018 9:12
 Operator : SM/SJ
 Sample : J5270-03RE
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 STOCK-PILE-BRE

Manual Integrations
 APPROVED

Sohil
 10/11/2018 3:00:10 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 09:47:10 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 08 15:27:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.393	3.514	2040159	2538372	9.025m	10.524m
2) SA Decachlor...	10.107	8.441	1519973	1314002	4.562m	5.024

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101018\
Data File : PO049850.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Oct 2018 9:12
Operator : SM/SJ
Sample : J5270-03RE
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
STOCK-PILE-BRE


Manual Integrations
APPROVED

Sohil
10/11/2018 3:00:10 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 09:47:10 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100818.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 08 15:27:38 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.393 min
Delta R.T.: 0.011 min
Response: 2040159
Conc: 9.03 ng/ml m

Instrument :
ECD_O
ClientSampleId :
STOCK-PILE-BRE


Manual Integrations
APPROVED

Sohil
10/11/2018 3:00:10 PM




#1 Tetrachloro-m-xylene

R.T.: 3.514 min
Delta R.T.: 0.000 min
Response: 2538372
Conc: 10.52 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.107 min
Delta R.T.: 0.020 min
Response: 1519973
Conc: 4.56 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.441 min
Delta R.T.: 0.004 min
Response: 1314002
Conc: 5.02 ng/ml