

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101019\
 Data File : PO061866.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Oct 2019 17:13
 Operator : HP/AJ
 Sample : K5297-04
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 03:14:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 10 11:46:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.352	3.513	993144	750210	24.215	23.198
2)	SA Decachlor...	10.010	8.373	759909	507268	16.317	15.668
Target Compounds							
3)	L1 AR-1016-1	5.540	4.580	11729	39981	6.396	29.733 #
4)	L1 AR-1016-2	5.572	4.615	2180	15590	0.843	8.412 #
5)	L1 AR-1016-3	5.637	4.770	2912	7055	1.813	7.043 #
6)	L1 AR-1016-4	5.736	4.834	12508	32402	9.355	37.392 #
7)	L1 AR-1016-5	6.033	5.027	8687	31206	6.334	27.339 #
8)	L2 AR-1221-1	4.576	3.727	6716	63472	14.528	160.292 #
9)	L2 AR-1221-2	4.656	3.801	27749	69668	77.291	225.906 #
10)	L2 AR-1221-3	4.749	3.893	46016	24051	40.418	26.367 #
11)	L3 AR-1232-1	4.749	3.893	46016	24051	32.906	21.420 #
12)	L3 AR-1232-2	5.252	4.603	56622	8995	72.170	7.369 #
13)	L3 AR-1232-3	5.540	4.761	11729	8980	6.864	13.600 #
14)	L3 AR-1232-4	5.719	4.852	4583	34458	5.180	54.665 #
15)	L3 AR-1232-5	5.790	5.027	18520	31206	26.711	45.479 #
16)	L4 AR-1242-1	5.540	4.568	11729	32570	8.259	31.791 #
17)	L4 AR-1242-2	5.540	4.580	11729	39981	5.836	28.351 #
18)	L4 AR-1242-3	5.618	4.761	3761	8980	2.980	11.499 #
19)	L4 AR-1242-4	5.719	4.834	4583	32402	4.361	39.486 #
20)	L4 AR-1242-5	6.435	5.344	79467	66484	59.415	62.062
21)	L5 AR-1248-1	5.540	4.580	11729	39981	9.984	44.681 #
22)	L5 AR-1248-2	5.833	4.834	9040	32402	5.358	27.500 #
23)	L5 AR-1248-3	6.033	4.852	8687	34458	4.648	27.760 #
24)	L5 AR-1248-4	6.435	5.027	79467	31206	33.632	20.758 #
25)	L5 AR-1248-5	6.435	5.411	79467	31664	35.047	20.639 #
26)	L6 AR-1254-1	6.395	5.344	43506	66484	19.070	30.312 #
27)	L6 AR-1254-2	6.585	5.507	164553	30640	45.544	15.629 #
28)	L6 AR-1254-3	6.953	5.877	129318	149463	33.096	46.945 #
29)	L6 AR-1254-4	7.235	6.132	127626	45146	42.140	22.252 #
30)	L6 AR-1254-5	7.652	6.545	133329	35273	42.456	12.471 #
31)	L7 AR-1260-1	7.177	6.052	71506	57216	27.391	28.050
32)	L7 AR-1260-2	7.413	6.221	120908	42164	37.926	17.556 #
33)	L7 AR-1260-3	7.783	6.338	87889	80325	36.592	35.651
34)	L7 AR-1260-4	7.999	6.838	30386	45727	11.372	24.768 #
35)	L7 AR-1260-5	8.295	7.081	20438	11534	4.131	2.962 #
36)	L8 AR-1262-1	7.726	6.601	75652	41335	19.745	33.864 #
37)	L8 AR-1262-2	8.265	7.081	87249	11534	14.054	2.508 #
38)	L8 AR-1262-3	8.585	7.352	120055	5529	28.885	2.935 #
39)	L8 AR-1262-4	8.660	7.382	30999	30820	9.820	9.171
40)	L8 AR-1262-5	9.292	7.899	40703	4016	20.015	2.649 #
41)	L9 AR-1268-1	8.585	7.352	120055	5529	17.854	1.145 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101019\
Data File : P0061866.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Oct 2019 17:13
Operator : HP/AJ
Sample : K5297-04
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 03:14:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 10 11:46:12 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.682	7.382	11959	30820	1.931	7.151 #
43)	L9 AR-1268-3	8.869	7.611	44327	2997	8.671	0.836 #
44)	L9 AR-1268-4	9.292	7.899	40703	4016	18.615	2.496 #
45)	L9 AR-1268-5	9.730	8.182	1067	2054	0.071	0.204 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101019\
Data File : P0061866.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Oct 2019 17:13
Operator : HP/AJ
Sample : K5297-04
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 03:14:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100119.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 10 11:46:12 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.352 min
Delta R.T.: -0.011 min
Response: 993144
Conc: 24.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#1 Tetrachloro-m-xylene

R.T.: 3.513 min
Delta R.T.: -0.006 min
Response: 750210
Conc: 23.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.010 min
Delta R.T.: -0.026 min
Response: 759909
Conc: 16.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.373 min
Delta R.T.: -0.036 min
Response: 507268
Conc: 15.67 ng/ml



#3 AR-1016-1


R.T.: 5.540 min
Delta R.T.: -0.010 min
Response: 11729
Conc: 6.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#3 AR-1016-1

R.T.: 4.580 min
Delta R.T.: -0.010 min
Response: 39981
Conc: 29.73 ng/ml




#4 AR-1016-2

R.T.: 5.572 min
Delta R.T.: 0.002 min
Response: 2180
Conc: 0.84 ng/ml



#4 AR-1016-2


R.T.: 4.615 min
Delta R.T.: 0.005 min
Response: 15590
Conc: 8.41 ng/ml



#5 AR-1016-3


R.T.: 5.637 min
Delta R.T.: -0.003 min
Response: 2912
Conc: 1.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#5 AR-1016-3

R.T.: 4.770 min
Delta R.T.: -0.010 min
Response: 7055
Conc: 7.04 ng/ml




#6 AR-1016-4

R.T.: 5.736 min
Delta R.T.: 0.006 min
Response: 12508
Conc: 9.36 ng/ml



#6 AR-1016-4


R.T.: 4.834 min
Delta R.T.: 0.014 min
Response: 32402
Conc: 37.39 ng/ml



#7 AR-1016-5


R.T.: 6.033 min
Delta R.T.: 0.003 min
Response: 8687
Conc: 6.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#7 AR-1016-5

R.T.: 5.027 min
Delta R.T.: -0.003 min
Response: 31206
Conc: 27.34 ng/ml




#8 AR-1221-1

R.T.: 4.576 min
Delta R.T.: 0.002 min
Response: 6716
Conc: 14.53 ng/ml



#8 AR-1221-1


R.T.: 3.727 min
Delta R.T.: -0.008 min
Response: 63472
Conc: 160.29 ng/ml



#9 AR-1221-2


R.T.: 4.656 min
Delta R.T.: -0.004 min
Response: 27749
Conc: 77.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#9 AR-1221-2

R.T.: 3.801 min
Delta R.T.: -0.016 min
Response: 69668
Conc: 225.91 ng/ml




#10 AR-1221-3

R.T.: 4.749 min
Delta R.T.: 0.012 min
Response: 46016
Conc: 40.42 ng/ml



#10 AR-1221-3


R.T.: 3.893 min
Delta R.T.: 0.002 min
Response: 24051
Conc: 26.37 ng/ml



#11 AR-1232-1


R.T.: 4.749 min
Delta R.T.: 0.013 min
Response: 46016
Conc: 32.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#11 AR-1232-1

R.T.: 3.893 min
Delta R.T.: 0.003 min
Response: 24051
Conc: 21.42 ng/ml




#12 AR-1232-2

R.T.: 5.252 min
Delta R.T.: -0.009 min
Response: 56622
Conc: 72.17 ng/ml



#12 AR-1232-2


R.T.: 4.603 min
Delta R.T.: 0.008 min
Response: 8995
Conc: 7.37 ng/ml



#13 AR-1232-3


R.T.: 5.540 min
Delta R.T.: -0.010 min
Response: 11729
Conc: 6.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#13 AR-1232-3

R.T.: 4.761 min
Delta R.T.: -0.004 min
Response: 8980
Conc: 13.60 ng/ml




#14 AR-1232-4

R.T.: 5.719 min
Delta R.T.: 0.009 min
Response: 4583
Conc: 5.18 ng/ml



#14 AR-1232-4


R.T.: 4.852 min
Delta R.T.: 0.005 min
Response: 34458
Conc: 54.67 ng/ml



#15 AR-1232-5


R.T.: 5.790 min
Delta R.T.: -0.009 min
Response: 18520
Conc: 26.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#15 AR-1232-5

R.T.: 5.027 min
Delta R.T.: 0.014 min
Response: 31206
Conc: 45.48 ng/ml




#16 AR-1242-1

R.T.: 5.540 min
Delta R.T.: 0.010 min
Response: 11729
Conc: 8.26 ng/ml



#16 AR-1242-1


R.T.: 4.568 min
Delta R.T.: -0.002 min
Response: 32570
Conc: 31.79 ng/ml



#17 AR-1242-2


R.T.: 5.540 min
Delta R.T.: -0.010 min
Response: 11729
Conc: 5.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#17 AR-1242-2

R.T.: 4.580 min
Delta R.T.: -0.010 min
Response: 39981
Conc: 28.35 ng/ml




#18 AR-1242-3

R.T.: 5.618 min
Delta R.T.: 0.008 min
Response: 3761
Conc: 2.98 ng/ml



#18 AR-1242-3


R.T.: 4.761 min
Delta R.T.: 0.001 min
Response: 8980
Conc: 11.50 ng/ml



#19 AR-1242-4


R.T.: 5.719 min
Delta R.T.: 0.009 min
Response: 4583
Conc: 4.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#19 AR-1242-4

R.T.: 4.834 min
Delta R.T.: -0.006 min
Response: 32402
Conc: 39.49 ng/ml




#20 AR-1242-5

R.T.: 6.435 min
Delta R.T.: -0.005 min
Response: 79467
Conc: 59.41 ng/ml



#20 AR-1242-5


R.T.: 5.344 min
Delta R.T.: -0.006 min
Response: 66484
Conc: 62.06 ng/ml



#21 AR-1248-1


R.T.: 5.540 min
Delta R.T.: -0.010 min
Response: 11729
Conc: 9.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#21 AR-1248-1

R.T.: 4.580 min
Delta R.T.: -0.010 min
Response: 39981
Conc: 44.68 ng/ml




#22 AR-1248-2

R.T.: 5.833 min
Delta R.T.: 0.013 min
Response: 9040
Conc: 5.36 ng/ml



#22 AR-1248-2


R.T.: 4.834 min
Delta R.T.: 0.014 min
Response: 32402
Conc: 27.50 ng/ml



#23 AR-1248-3


R.T.: 6.033 min
Delta R.T.: 0.003 min
Response: 8687
Conc: 4.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#23 AR-1248-3

R.T.: 4.852 min
Delta R.T.: -0.008 min
Response: 34458
Conc: 27.76 ng/ml




#24 AR-1248-4

R.T.: 6.435 min
Delta R.T.: 0.005 min
Response: 79467
Conc: 33.63 ng/ml



#24 AR-1248-4


R.T.: 5.027 min
Delta R.T.: -0.003 min
Response: 31206
Conc: 20.76 ng/ml



#25 AR-1248-5


R.T.: 6.435 min
Delta R.T.: -0.035 min
Response: 79467
Conc: 35.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#25 AR-1248-5

R.T.: 5.411 min
Delta R.T.: 0.000 min
Response: 31664
Conc: 20.64 ng/ml




#26 AR-1254-1

R.T.: 6.395 min
Delta R.T.: 0.014 min
Response: 43506
Conc: 19.07 ng/ml



#26 AR-1254-1


R.T.: 5.344 min
Delta R.T.: -0.013 min
Response: 66484
Conc: 30.31 ng/ml



#27 AR-1254-2


R.T.: 6.585 min
Delta R.T.: -0.014 min
Response: 164553
Conc: 45.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.006 min
Response: 30640
Conc: 15.63 ng/ml




#28 AR-1254-3

R.T.: 6.953 min
Delta R.T.: -0.011 min
Response: 129318
Conc: 33.10 ng/ml



#28 AR-1254-3


R.T.: 5.877 min
Delta R.T.: -0.018 min
Response: 149463
Conc: 46.95 ng/ml



#29 AR-1254-4


R.T.: 7.235 min
Delta R.T.: -0.015 min
Response: 127626
Conc: 42.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#29 AR-1254-4

R.T.: 6.132 min
Delta R.T.: 0.013 min
Response: 45146
Conc: 22.25 ng/ml




#30 AR-1254-5

R.T.: 7.652 min
Delta R.T.: -0.015 min
Response: 133329
Conc: 42.46 ng/ml



#30 AR-1254-5


R.T.: 6.545 min
Delta R.T.: 0.015 min
Response: 35273
Conc: 12.47 ng/ml



#31 AR-1260-1


R.T.: 7.177 min
Delta R.T.: 0.027 min
Response: 71506
Conc: 27.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#31 AR-1260-1

R.T.: 6.052 min
Delta R.T.: 0.012 min
Response: 57216
Conc: 28.05 ng/ml




#32 AR-1260-2

R.T.: 7.413 min
Delta R.T.: 0.013 min
Response: 120908
Conc: 37.93 ng/ml



#32 AR-1260-2


R.T.: 6.221 min
Delta R.T.: 0.001 min
Response: 42164
Conc: 17.56 ng/ml



#33 AR-1260-3


R.T.: 7.783 min
Delta R.T.: 0.023 min
Response: 87889
Conc: 36.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#33 AR-1260-3

R.T.: 6.338 min
Delta R.T.: -0.032 min
Response: 80325
Conc: 35.65 ng/ml




#34 AR-1260-4

R.T.: 7.999 min
Delta R.T.: 0.009 min
Response: 30386
Conc: 11.37 ng/ml



#34 AR-1260-4


R.T.: 6.838 min
Delta R.T.: -0.002 min
Response: 45727
Conc: 24.77 ng/ml



#35 AR-1260-5


R.T.: 8.295 min
Delta R.T.: -0.005 min
Response: 20438
Conc: 4.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#35 AR-1260-5

R.T.: 7.081 min
Delta R.T.: 0.011 min
Response: 11534
Conc: 2.96 ng/ml




#36 AR-1262-1

R.T.: 7.726 min
Delta R.T.: -0.013 min
Response: 75652
Conc: 19.74 ng/ml



#36 AR-1262-1


R.T.: 6.601 min
Delta R.T.: -0.018 min
Response: 41335
Conc: 33.86 ng/ml



#37 AR-1262-2


R.T.: 8.265 min
Delta R.T.: -0.014 min
Response: 87249
Conc: 14.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#37 AR-1262-2

R.T.: 7.081 min
Delta R.T.: 0.020 min
Response: 11534
Conc: 2.51 ng/ml




#38 AR-1262-3

R.T.: 8.585 min
Delta R.T.: 0.003 min
Response: 120055
Conc: 28.88 ng/ml



#38 AR-1262-3


R.T.: 7.352 min
Delta R.T.: 0.012 min
Response: 5529
Conc: 2.93 ng/ml



#39 AR-1262-4


R.T.: 8.660 min
Delta R.T.: -0.009 min
Response: 30999
Conc: 9.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#39 AR-1262-4

R.T.: 7.382 min
Delta R.T.: -0.021 min
Response: 30820
Conc: 9.17 ng/ml




#40 AR-1262-5

R.T.: 9.292 min
Delta R.T.: -0.015 min
Response: 40703
Conc: 20.01 ng/ml



#40 AR-1262-5


R.T.: 7.899 min
Delta R.T.: 0.008 min
Response: 4016
Conc: 2.65 ng/ml



#41 AR-1268-1


R.T.: 8.585 min
Delta R.T.: 0.003 min
Response: 120055
Conc: 17.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#41 AR-1268-1

R.T.: 7.352 min
Delta R.T.: 0.011 min
Response: 5529
Conc: 1.14 ng/ml




#42 AR-1268-2

R.T.: 8.682 min
Delta R.T.: 0.007 min
Response: 11959
Conc: 1.93 ng/ml



#42 AR-1268-2


R.T.: 7.382 min
Delta R.T.: -0.022 min
Response: 30820
Conc: 7.15 ng/ml



#43 AR-1268-3


R.T.: 8.869 min
Delta R.T.: -0.025 min
Response: 44327
Conc: 8.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1




#43 AR-1268-3

R.T.: 7.611 min
Delta R.T.: 0.004 min
Response: 2997
Conc: 0.84 ng/ml




#44 AR-1268-4

R.T.: 9.292 min
Delta R.T.: -0.017 min
Response: 40703
Conc: 18.61 ng/ml



#44 AR-1268-4


R.T.: 7.899 min
Delta R.T.: 0.008 min
Response: 4016
Conc: 2.50 ng/ml



#45 AR-1268-5

R.T.: 9.730 min
Delta R.T.: 0.022 min
Response: 1067
Conc: 0.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1



#45 AR-1268-5

R.T.: 8.182 min
Delta R.T.: 0.019 min
Response: 2054
Conc: 0.20 ng/ml