

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101024\
 Data File : P0107091.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2024 01:20
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 05:40:40 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 04:48:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.373	3.646	572.4E6	206.0E6	47.715	51.656
2) SA Decachlor...	10.051	8.641	128.3E6	124.1E6	51.273	44.208
Target Compounds						
26) L6 AR-1254-1	6.374	5.529	125.9E6	97342624	413.214	576.037 #
27) L6 AR-1254-2	6.592	5.676	213.6E6	81724256	502.382	543.343
28) L6 AR-1254-3	6.955	6.078	199.0E6	119.7E6	502.207	513.907
29) L6 AR-1254-4	7.239	6.306	137.7E6	69994005	519.142	547.688
30) L6 AR-1254-5	7.659	6.724	116.9E6	97498700	418.560	484.622

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101024\
Data File : PO107091.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Oct 2024 01:20
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 05:40:40 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 04:48:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

