

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101024\
 Data File : P0107056.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2024 13:07
 Operator : YP/AJ
 Sample : PB164017BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB164017BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 01:47:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 04:48:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.374	3.646	198.9E6	63534954	16.582	15.929
2) SA Decachlor...	10.052	8.642	49044099	49514768	19.600	17.640

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101024\
Data File : PO107056.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2024 13:07
Operator : YP/AJ
Sample : PB164017BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB164017BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:47:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 04:48:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.374 min
Delta R.T.: -0.002 min
Response: 198919508
Conc: 16.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB164017BL




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: -0.002 min
Response: 63534954
Conc: 15.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.052 min
Delta R.T.: -0.004 min
Response: 49044099
Conc: 19.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.642 min
Delta R.T.: -0.005 min
Response: 49514768
Conc: 17.64 ng/ml