

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101119\
 Data File : P0061913.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Oct 2019 9:43
 Operator : HP/AJ
 Sample : K5316-03
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-9

Manual Integrations
 APPROVED

Ankita
 10/14/2019 1:08:04 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 00:47:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 12 06:25:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.325	3.482	1094730	808453	17.701	18.853
2)	SA Decachlor...	9.953	8.337	665844	329759	11.900	11.239
Target Compounds							
31)	L7 AR-1260-1	7.079	5.968	2544893	1645640	893.846	774.515
32)	L7 AR-1260-2	7.333	6.154	4637064	1829609	1351.696	720.993 #
33)	L7 AR-1260-3	7.692	6.304	3486758	1647170	1317.689	729.448 #
34)	L7 AR-1260-4	7.918	6.766	3137728	1945016	1121.800	1107.735
35)	L7 AR-1260-5	8.228	7.006	7506152	4509883	1433.675m	1251.897

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101119\
Data File : PO061913.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Oct 2019 9:43
Operator : HP/AJ
Sample : K5316-03
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
TP-9


Manual Integrations
APPROVED

Ankita
10/14/2019 1:08:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 00:47:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 1094730
Conc: 17.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-9


Manual Integrations
APPROVED

Ankita
10/14/2019 1:08:04 PM




#1 Tetrachloro-m-xylene

R.T.: 3.482 min
Delta R.T.: 0.000 min
Response: 808453
Conc: 18.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.953 min
Delta R.T.: -0.003 min
Response: 665844
Conc: 11.90 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.337 min
Delta R.T.: -0.001 min
Response: 329759
Conc: 11.24 ng/ml




#31 AR-1260-1

R.T.: 7.079 min
Delta R.T.: 0.000 min
Response: 2544893
Conc: 893.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-9


Manual Integrations
APPROVED

Ankita
10/14/2019 1:08:04 PM




#31 AR-1260-1

R.T.: 5.968 min
Delta R.T.: 0.000 min
Response: 1645640
Conc: 774.52 ng/ml




#32 AR-1260-2

R.T.: 7.333 min
Delta R.T.: -0.003 min
Response: 4637064
Conc: 1351.70 ng/ml



#32 AR-1260-2

R.T.: 6.154 min
Delta R.T.: 0.000 min
Response: 1829609
Conc: 720.99 ng/ml




#33 AR-1260-3

R.T.: 7.692 min
Delta R.T.: -0.002 min
Response: 3486758
Conc: 1317.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-9


Manual Integrations
APPROVED

Ankita
10/14/2019 1:08:04 PM




#33 AR-1260-3

R.T.: 6.304 min
Delta R.T.: 0.000 min
Response: 1647170
Conc: 729.45 ng/ml




#34 AR-1260-4

R.T.: 7.918 min
Delta R.T.: 0.000 min
Response: 3137728
Conc: 1121.80 ng/ml



#34 AR-1260-4

R.T.: 6.766 min
Delta R.T.: 0.000 min
Response: 1945016
Conc: 1107.74 ng/ml




#35 AR-1260-5

R.T.: 8.228 min
Delta R.T.: -0.002 min
Response: 7506152
Conc: 1433.67 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TP-9

Manual Integrations
APPROVED

Ankita
10/14/2019 1:08:04 PM



#35 AR-1260-5

R.T.: 7.006 min
Delta R.T.: 0.000 min
Response: 4509883
Conc: 1251.90 ng/ml