

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101121\
 Data File : PO081948.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Oct 2021 9:27
 Operator : AJ\MA
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 16:29:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 07 09:43:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
2)	SA Decachlor...	11.130f	0.000	21601	0	0.364	N.D. #
Target Compounds							
8)	L2 AR-1221-1	5.243	0.000	51424	0	53.848	N.D. #
9)	L2 AR-1221-2	5.348	0.000	19385	0	30.026	N.D. #
10)	L2 AR-1221-3	5.416f	4.423	6993	20949	3.279	27.780 #
11)	L3 AR-1232-1	5.416f	4.423	6993	20949	4.192	35.322 #
30)	L6 AR-1254-5	8.563f	0.000	27618	0	8.459	N.D. #
33)	L7 AR-1260-3	0.000	7.192	0	137141	N.D.	83.860 #
39)	L8 AR-1262-4	9.619f	8.270	229297	54781	113.508	22.758 #
42)	L9 AR-1268-2	9.619f	8.270	229297	54781	29.132	15.476 #
45)	L9 AR-1268-5	10.767	0.000	23170	0	1.042	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101121\
Data File : PO081948.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Oct 2021 9:27
Operator : AJ\MA
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 16:29:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 07 09:43:36 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 4.990 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.981 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 11.130 min
Delta R.T.: 0.020 min
Response: 21601
Conc: 0.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 9.316 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 5.243 min
Delta R.T.: 0.000 min
Response: 51424
Conc: 53.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.243 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.348 min
Delta R.T.: 0.005 min
Response: 19385
Conc: 30.03 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.342 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 5.416 min
Delta R.T.: -0.016 min
Response: 6993
Conc: 3.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#10 AR-1221-3

R.T.: 4.423 min
Delta R.T.: -0.008 min
Response: 20949
Conc: 27.78 ng/ml




#11 AR-1232-1

R.T.: 5.416 min
Delta R.T.: -0.015 min
Response: 6993
Conc: 4.19 ng/ml



#11 AR-1232-1


R.T.: 4.423 min
Delta R.T.: -0.007 min
Response: 20949
Conc: 35.32 ng/ml



#30 AR-1254-5


R.T.: 8.563 min
Delta R.T.: -0.021 min
Response: 27618
Conc: 8.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.375 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 8.658 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.192 min
Delta R.T.: -0.005 min
Response: 137141
Conc: 83.86 ng/ml



#39 AR-1262-4


R.T.: 9.619 min
Delta R.T.: -0.025 min
Response: 229297
Conc: 113.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 8.270 min
Delta R.T.: -0.013 min
Response: 54781
Conc: 22.76 ng/ml




#42 AR-1268-2

R.T.: 9.619 min
Delta R.T.: -0.024 min
Response: 229297
Conc: 29.13 ng/ml



#42 AR-1268-2


R.T.: 8.270 min
Delta R.T.: -0.012 min
Response: 54781
Conc: 15.48 ng/ml



#45 AR-1268-5

R.T.: 10.767 min
Delta R.T.: 0.008 min
Response: 23170
Conc: 1.04 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 9.065 min
Response: 0
Conc: N.D.