

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101121\
 Data File : PO081954.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Oct 2021 11:11
 Operator : AJ\MA
 Sample : PB139777BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 16:31:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 07 09:43:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.989	3.978	1603228	587197	19.320	23.205
2) SA Decachlor...	11.109	9.304	1229044	574151	20.721	24.305
Target Compounds						
8) L2 AR-1221-1	5.241	0.000	200258	0	209.698	N.D. #
9) L2 AR-1221-2	5.344	0.000	21103	0	32.687	N.D. #
10) L2 AR-1221-3	0.000	4.426	0	18804	N.D.	24.936 #
11) L3 AR-1232-1	0.000	4.426	0	18804	N.D.	31.706 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101121\
Data File : PO081954.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Oct 2021 11:11
Operator : AJ\MA
Sample : PB139777BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 16:31:36 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100621.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 07 09:43:36 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.989 min
Delta R.T.: 0.000 min
Response: 1603228
Conc: 19.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.978 min
Delta R.T.: -0.004 min
Response: 587197
Conc: 23.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.109 min
Delta R.T.: 0.000 min
Response: 1229044
Conc: 20.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.304 min
Delta R.T.: -0.012 min
Response: 574151
Conc: 24.30 ng/ml



#8 AR-1221-1


R.T.: 5.241 min
Delta R.T.: -0.003 min
Response: 200258
Conc: 209.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.243 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.344 min
Delta R.T.: 0.000 min
Response: 21103
Conc: 32.69 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.342 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 5.431 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#10 AR-1221-3

R.T.: 4.426 min
Delta R.T.: -0.005 min
Response: 18804
Conc: 24.94 ng/ml



#11 AR-1232-1

R.T.: 0.000 min
Exp R.T.: 5.431 min
Response: 0
Conc: N.D.



#11 AR-1232-1

R.T.: 4.426 min
Delta R.T.: -0.004 min
Response: 18804
Conc: 31.71 ng/ml