

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101123\
 Data File : PO098716.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2023 02:45
 Operator : YP/AJ
 Sample : 04813-06DL 5X
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-20DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/12/2023
 Supervised By :Ankita Jodhani 10/12/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 03:33:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.467	3.633	7050875	3890268	3.672m	4.995 #
2) SA Decachlor...	10.291	8.639	6556740	2763197	5.402	5.157
Target Compounds						
16) L4 AR-1242-1	5.654	4.716	405.2E6	156.7E6	8899.024	7486.279
17) L4 AR-1242-2	5.676	4.734	597.9E6	220.4E6	8730.283	7579.317
18) L4 AR-1242-3	5.738	4.910	379.4E6	121.5E6	8960.077	7776.655
19) L4 AR-1242-4	5.838	4.995	315.4E6	94554695	9611.030	5709.012 #
20) L4 AR-1242-5	6.578	5.518	228.0E6	116.0E6	6436.604	6251.957
31) L7 AR-1260-1	7.264	6.200	52674214	28964715	724.393	789.432
32) L7 AR-1260-2	7.522	6.388	49525696	21395809	601.472	509.904
33) L7 AR-1260-3	7.882	6.542	29821205	22791580	482.056	569.546
34) L7 AR-1260-4	8.109	7.013	34103353	12800276	505.890	397.486
35) L7 AR-1260-5	8.434	7.256	35890166	17886117	276.249	255.296

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101123\
Data File : PO098716.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2023 02:45
Operator : YP/AJ
Sample : 04813-06DL 5X
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB-20DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/12/2023
Supervised By :Ankita Jodhani 10/12/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 03:33:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

