

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101123\
 Data File : PO098669.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2023 11:29
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 14:46:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 28 01:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.479	3.634	94903666	38895486	49.424	49.939
2) SA Decachlor...	10.290	8.641	60361015	27795979	49.726	51.881
Target Compounds						
21) L5 AR-1248-1	5.654	4.717	16910890	7599224	499.689	484.917
22) L5 AR-1248-2	5.929	4.953	25293100	11088628	488.389	503.875
23) L5 AR-1248-3	6.133	4.995	27727322	11653904	490.634	496.338
24) L5 AR-1248-4	6.539	5.167	27495650	13168960	485.030	493.695
25) L5 AR-1248-5	6.578	5.560	28613448	11526860	483.761	488.495



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101123\
Data File : PO098669.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Oct 2023 11:29
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 14:46:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO092723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 28 01:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: -0.006 min
Response: 94903666
Conc: 49.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.634 min
Delta R.T.: -0.004 min
Response: 38895486
Conc: 49.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.290 min
Delta R.T.: -0.003 min
Response: 60361015
Conc: 49.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.641 min
Delta R.T.: -0.011 min
Response: 27795979
Conc: 51.88 ng/ml



#21 AR-1248-1


R.T.: 5.654 min
Delta R.T.: -0.009 min
Response: 16910890
Conc: 499.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.717 min
Delta R.T.: -0.005 min
Response: 7599224
Conc: 484.92 ng/ml




#22 AR-1248-2

R.T.: 5.929 min
Delta R.T.: -0.008 min
Response: 25293100
Conc: 488.39 ng/ml



#22 AR-1248-2


R.T.: 4.953 min
Delta R.T.: -0.005 min
Response: 11088628
Conc: 503.88 ng/ml



#23 AR-1248-3


R.T.: 6.133 min
Delta R.T.: -0.009 min
Response: 27727322
Conc: 490.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.995 min
Delta R.T.: -0.006 min
Response: 11653904
Conc: 496.34 ng/ml




#24 AR-1248-4

R.T.: 6.539 min
Delta R.T.: -0.010 min
Response: 27495650
Conc: 485.03 ng/ml



#24 AR-1248-4


R.T.: 5.167 min
Delta R.T.: -0.006 min
Response: 13168960
Conc: 493.70 ng/ml



#25 AR-1248-5

R.T.: 6.578 min
Delta R.T.: -0.009 min
Response: 28613448
Conc: 483.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.560 min
Delta R.T.: -0.007 min
Response: 11526860
Conc: 488.49 ng/ml