

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101220\
 Data File : PO072280.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Oct 2020 14:34
 Operator : DD\AJ
 Sample : L4337-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-4

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 15:55:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 05 11:48:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.325	3.510	1768411	920938	24.621	24.587
2) SA Decachlor...	9.928	8.677	1633688	828740	14.837	14.570

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101220\
Data File : PO072280.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Oct 2020 14:34
Operator : DD\AJ
Sample : L4337-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-4

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 15:55:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100320.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 05 11:48:20 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.325 min
Delta R.T.: -0.004 min
Response: 1768411
Conc: 24.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-4




#1 Tetrachloro-m-xylene

R.T.: 3.510 min
Delta R.T.: -0.003 min
Response: 920938
Conc: 24.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.928 min
Delta R.T.: -0.009 min
Response: 1633688
Conc: 14.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.677 min
Delta R.T.: -0.010 min
Response: 828740
Conc: 14.57 ng/ml