

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101320\
 Data File : PO072313.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Oct 2020 15:35
 Operator : DD\AJ
 Sample : L4369-04
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HD-2

Manual Integrations
 APPROVED

Ankita
 10/14/2020 12:02:34 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 02:23:05 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 05 11:48:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.325	3.509	1887297	940830	26.276	25.118
2) SA Decachlor...	9.931	8.679	1594210	749334	14.479	13.174
Target Compounds						
26) L6 AR-1254-1	6.518	5.580	1598225	881702	457.453	368.938
27) L6 AR-1254-2	6.746	5.742	2559243	894408	464.850	418.149
28) L6 AR-1254-3	7.123	6.151	3456285	2090781	598.473	615.813
29) L6 AR-1254-4	7.417	6.391	2327106	871038	399.438	358.721
30) L6 AR-1254-5	7.838	6.812	4597252	2650216	854.212m	809.941

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101320\
Data File : PO072313.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Oct 2020 15:35
Operator : DD\AJ
Sample : L4369-04
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
HD-2


Manual Integrations
APPROVED

Ankita
10/14/2020 12:02:34 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 02:23:05 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100320.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 05 11:48:20 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.325 min
Delta R.T.: -0.004 min
Response: 1887297
Conc: 26.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
HD-2


Manual Integrations
APPROVED

Ankita
10/14/2020 12:02:34 PM




#1 Tetrachloro-m-xylene

R.T.: 3.509 min
Delta R.T.: -0.004 min
Response: 940830
Conc: 25.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.931 min
Delta R.T.: -0.006 min
Response: 1594210
Conc: 14.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.679 min
Delta R.T.: -0.008 min
Response: 749334
Conc: 13.17 ng/ml




#26 AR-1254-1

R.T.: 6.518 min
Delta R.T.: -0.003 min
Response: 1598225
Conc: 457.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
HD-2


Manual Integrations
APPROVED

Ankita
10/14/2020 12:02:34 PM




#26 AR-1254-1

R.T.: 5.580 min
Delta R.T.: -0.004 min
Response: 881702
Conc: 368.94 ng/ml




#27 AR-1254-2

R.T.: 6.746 min
Delta R.T.: -0.002 min
Response: 2559243
Conc: 464.85 ng/ml



#27 AR-1254-2

R.T.: 5.742 min
Delta R.T.: -0.003 min
Response: 894408
Conc: 418.15 ng/ml




#28 AR-1254-3

R.T.: 7.123 min
Delta R.T.: -0.002 min
Response: 3456285
Conc: 598.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
HD-2


Manual Integrations
APPROVED

Ankita
10/14/2020 12:02:34 PM




#28 AR-1254-3

R.T.: 6.151 min
Delta R.T.: -0.003 min
Response: 2090781
Conc: 615.81 ng/ml




#29 AR-1254-4

R.T.: 7.417 min
Delta R.T.: -0.002 min
Response: 2327106
Conc: 399.44 ng/ml



#29 AR-1254-4

R.T.: 6.391 min
Delta R.T.: -0.003 min
Response: 871038
Conc: 358.72 ng/ml




#30 AR-1254-5

R.T.: 7.838 min
Delta R.T.: -0.002 min
Response: 4597252
Conc: 854.21 ng/ml m

Instrument :
ECD_O
ClientSampleId :
HD-2

Manual Integrations
APPROVED

Ankita
10/14/2020 12:02:34 PM



#30 AR-1254-5

R.T.: 6.812 min
Delta R.T.: -0.004 min
Response: 2650216
Conc: 809.94 ng/ml