

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101325\
 Data File : PO114351.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2025 11:07
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 02:30:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 06:46:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.357	3.661	303.3E6	230.8E6	55.869	57.680
2) SA Decachlor...	9.940	8.672	187.9E6	174.8E6	51.617	52.439
Target Compounds						
26) L6 AR-1254-1	6.338	5.544	91102755	138.2E6	522.218	520.031
27) L6 AR-1254-2	6.554	5.693	131.2E6	123.3E6	515.278	517.455
28) L6 AR-1254-3	6.913	6.095	144.9E6	184.6E6	501.649	520.743
29) L6 AR-1254-4	7.195	6.323	121.5E6	133.5E6	513.555	517.213
30) L6 AR-1254-5	7.609	6.741	114.0E6	165.4E6	519.548	516.669



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101325\
Data File : PO114351.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Oct 2025 11:07
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 02:30:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 06:46:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.357 min
Delta R.T.: 0.002 min
Response: 303311220
Conc: 55.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.661 min
Delta R.T.: 0.000 min
Response: 230848727
Conc: 57.68 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.940 min
Delta R.T.: 0.005 min
Response: 187941532
Conc: 51.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.672 min
Delta R.T.: 0.000 min
Response: 174787114
Conc: 52.44 ng/ml



#26 AR-1254-1


R.T.: 6.338 min
Delta R.T.: 0.003 min
Response: 91102755
Conc: 522.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.544 min
Delta R.T.: 0.002 min
Response: 138169067
Conc: 520.03 ng/ml




#27 AR-1254-2

R.T.: 6.554 min
Delta R.T.: 0.004 min
Response: 131244650
Conc: 515.28 ng/ml



#27 AR-1254-2


R.T.: 5.693 min
Delta R.T.: 0.003 min
Response: 123274281
Conc: 517.45 ng/ml



#28 AR-1254-3


R.T.: 6.913 min
Delta R.T.: 0.004 min
Response: 144946635
Conc: 501.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.095 min
Delta R.T.: 0.003 min
Response: 184555239
Conc: 520.74 ng/ml




#29 AR-1254-4

R.T.: 7.195 min
Delta R.T.: 0.005 min
Response: 121530991
Conc: 513.56 ng/ml



#29 AR-1254-4


R.T.: 6.323 min
Delta R.T.: 0.004 min
Response: 133522016
Conc: 517.21 ng/ml



#30 AR-1254-5

R.T.: 7.609 min
Delta R.T.: 0.003 min
Response: 114028299
Conc: 519.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.741 min
Delta R.T.: 0.004 min
Response: 165439165
Conc: 516.67 ng/ml