

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101325\
 Data File : PO114360.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2025 15:45
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 02:33:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 06:46:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.357	3.661	308.1E6	227.9E6	56.760	56.941
2) SA Decachlor...	9.941	8.673	189.0E6	165.3E6	51.902	49.606
Target Compounds						
16) L4 AR-1242-1	5.495	4.743	69845539	68901368	530.888	527.524
17) L4 AR-1242-2	5.516	4.763	116.0E6	98218790	528.381	525.833
18) L4 AR-1242-3	5.578	4.938	66021298	52306412	535.324	546.542
19) L4 AR-1242-4	5.674	5.022	54104833	49411093	527.366	535.078
20) L4 AR-1242-5	6.400	5.543	55659974	70060877	480.515	571.643



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101325\
Data File : PO114360.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Oct 2025 15:45
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 02:33:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 06:46:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.357 min
Delta R.T.: 0.002 min
Response: 308147596
Conc: 56.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.661 min
Delta R.T.: 0.000 min
Response: 227894050
Conc: 56.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.941 min
Delta R.T.: 0.006 min
Response: 188980903
Conc: 51.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.673 min
Delta R.T.: 0.001 min
Response: 165345254
Conc: 49.61 ng/ml



#16 AR-1242-1


R.T.: 5.495 min
Delta R.T.: 0.002 min
Response: 69845539
Conc: 530.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.743 min
Delta R.T.: 0.001 min
Response: 68901368
Conc: 527.52 ng/ml




#17 AR-1242-2

R.T.: 5.516 min
Delta R.T.: 0.000 min
Response: 116007102
Conc: 528.38 ng/ml



#17 AR-1242-2


R.T.: 4.763 min
Delta R.T.: 0.001 min
Response: 98218790
Conc: 525.83 ng/ml



#18 AR-1242-3


R.T.: 5.578 min
Delta R.T.: 0.002 min
Response: 66021298
Conc: 535.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.938 min
Delta R.T.: 0.002 min
Response: 52306412
Conc: 546.54 ng/ml




#19 AR-1242-4

R.T.: 5.674 min
Delta R.T.: 0.002 min
Response: 54104833
Conc: 527.37 ng/ml



#19 AR-1242-4


R.T.: 5.022 min
Delta R.T.: 0.000 min
Response: 49411093
Conc: 535.08 ng/ml



#20 AR-1242-5

R.T.: 6.400 min
Delta R.T.: 0.003 min
Response: 55659974
Conc: 480.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.543 min
Delta R.T.: 0.000 min
Response: 70060877
Conc: 571.64 ng/ml