

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101424\
 Data File : P0107101.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Oct 2024 12:22
 Operator : YP/AJ
 Sample : P4368-07
 Misc : AR1248 LOD 25 PPB
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2024

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/15/2024
 Supervised By :Ankita Jodhani 10/15/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 13:30:33 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 04:48:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.374	3.646	171.3E6	55397944	14.282	13.889
2) SA Decachlor...	10.060	8.641	46967705	45630113	18.771	16.256
Target Compounds						
21) L5 AR-1248-1	5.522	4.728	5274689	1792209	26.785	26.019
22) L5 AR-1248-2	5.796	4.964	7787033	2588434	27.647	26.704
23) L5 AR-1248-3	5.999	5.006	7570064	2695569	26.280	26.613
24) L5 AR-1248-4	6.399	5.177	8041665	2764757	28.338	23.334
25) L5 AR-1248-5	6.438	5.568	8293516	2891990	30.028	27.029m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101424\
Data File : PO107101.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2024 12:22
Operator : YP/AJ
Sample : P4368-07
Misc : AR1248 LOD 25 PPB
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT4-2024

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/15/2024
Supervised By :Ankita Jodhani 10/15/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 13:30:33 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 04:48:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

