

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101424\
 Data File : P0107138.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Oct 2024 23:33
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 03:40:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 04:48:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.373	3.644	509.2E6	173.7E6	42.443	43.540
2)	SA Decachlor...	10.058	8.638	135.4E6	142.6E6	54.123	50.812
Target Compounds							
3)	L1 AR-1016-1	5.522	4.726	155.2E6	57690124	447.273	474.261
4)	L1 AR-1016-2	5.544	4.744	226.8E6	81400899	451.706	481.433
5)	L1 AR-1016-3	5.607	4.921	144.1E6	47321242	468.129	519.973
6)	L1 AR-1016-4	5.704	4.962	114.9E6	36544311	465.287	498.136
7)	L1 AR-1016-5	5.997	5.175	105.5E6	44611409	470.936	485.472
31)	L7 AR-1260-1	7.124	6.206	159.2E6	88230096	490.206	512.271
32)	L7 AR-1260-2	7.380	6.393	160.8E6	102.7E6	472.289	508.326
33)	L7 AR-1260-3	7.742	6.547	94315605	94808420	442.902	501.300
34)	L7 AR-1260-4	7.967	7.018	96565911	68912186	427.437	509.439
35)	L7 AR-1260-5	8.280	7.258	144.8E6	162.5E6	436.314	526.121



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101424\
Data File : PO107138.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2024 23:33
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 03:40:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 04:48:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.373 min
Delta R.T.: -0.003 min
Response: 509166139
Conc: 42.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: -0.004 min
Response: 173665172
Conc: 43.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.058 min
Delta R.T.: 0.002 min
Response: 135425792
Conc: 54.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.638 min
Delta R.T.: -0.010 min
Response: 142631007
Conc: 50.81 ng/ml



#3 AR-1016-1


R.T.: 5.522 min
Delta R.T.: -0.002 min
Response: 155207770
Conc: 447.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.726 min
Delta R.T.: -0.005 min
Response: 57690124
Conc: 474.26 ng/ml




#4 AR-1016-2

R.T.: 5.544 min
Delta R.T.: -0.002 min
Response: 226759359
Conc: 451.71 ng/ml



#4 AR-1016-2


R.T.: 4.744 min
Delta R.T.: -0.005 min
Response: 81400899
Conc: 481.43 ng/ml



#5 AR-1016-3


R.T.: 5.607 min
Delta R.T.: -0.002 min
Response: 144127553
Conc: 468.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.921 min
Delta R.T.: -0.005 min
Response: 47321242
Conc: 519.97 ng/ml




#6 AR-1016-4

R.T.: 5.704 min
Delta R.T.: -0.001 min
Response: 114873740
Conc: 465.29 ng/ml



#6 AR-1016-4


R.T.: 4.962 min
Delta R.T.: -0.005 min
Response: 36544311
Conc: 498.14 ng/ml



#7 AR-1016-5


R.T.: 5.997 min
Delta R.T.: -0.002 min
Response: 105475478
Conc: 470.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.175 min
Delta R.T.: -0.005 min
Response: 44611409
Conc: 485.47 ng/ml




#31 AR-1260-1

R.T.: 7.124 min
Delta R.T.: 0.000 min
Response: 159202149
Conc: 490.21 ng/ml



#31 AR-1260-1


R.T.: 6.206 min
Delta R.T.: -0.007 min
Response: 88230096
Conc: 512.27 ng/ml



#32 AR-1260-2


R.T.: 7.380 min
Delta R.T.: 0.000 min
Response: 160814288
Conc: 472.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.393 min
Delta R.T.: -0.006 min
Response: 102691210
Conc: 508.33 ng/ml




#33 AR-1260-3

R.T.: 7.742 min
Delta R.T.: 0.000 min
Response: 94315605
Conc: 442.90 ng/ml



#33 AR-1260-3


R.T.: 6.547 min
Delta R.T.: -0.006 min
Response: 94808420
Conc: 501.30 ng/ml



#34 AR-1260-4


R.T.: 7.967 min
Delta R.T.: 0.002 min
Response: 96565911
Conc: 427.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.018 min
Delta R.T.: -0.006 min
Response: 68912186
Conc: 509.44 ng/ml



#35 AR-1260-5

R.T.: 8.280 min
Delta R.T.: 0.001 min
Response: 144802028
Conc: 436.31 ng/ml



#35 AR-1260-5

R.T.: 7.258 min
Delta R.T.: -0.006 min
Response: 162509443
Conc: 526.12 ng/ml