

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101520\
 Data File : P0072376.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Oct 2020 13:09
 Operator : DD\AJ
 Sample : L4223-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PL-01-101420

Manual Integrations
 APPROVED

Ankita
 10/16/2020 11:53:35 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 14:01:22 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 14 17:47:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.325	3.506	1805629	876064	21.288m	20.798
2) SA Decachlor...	9.927	8.674	2238697	989907	20.296	19.149

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101520\
Data File : PO072376.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Oct 2020 13:09
Operator : DD\AJ
Sample : L4223-01
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PL-01-101420


Manual Integrations
APPROVED

Ankita
10/16/2020 11:53:35 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 14:01:22 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 14 17:47:24 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.325 min
Delta R.T.: 0.004 min
Response: 1805629
Conc: 21.29 ng/ml m

Instrument :
ECD_O
ClientSampleId :
PL-01-101420


Manual Integrations
APPROVED

Ankita
10/16/2020 11:53:35 AM




#1 Tetrachloro-m-xylene

R.T.: 3.506 min
Delta R.T.: 0.002 min
Response: 876064
Conc: 20.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.927 min
Delta R.T.: 0.001 min
Response: 2238697
Conc: 20.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.674 min
Delta R.T.: -0.001 min
Response: 989907
Conc: 19.15 ng/ml