

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101524\
 Data File : PO107213.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Oct 2024 03:12
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO101524AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 04:37:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 16 04:36:18 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.373	3.644	462.5E6	169.0E6	51.003	52.312
2)	SA Decachlor...	10.058	8.637	124.4E6	137.6E6	50.838	50.298
Target Compounds							
16)	L4 AR-1242-1	5.522	4.726	112.5E6	42391283	503.251	506.204
17)	L4 AR-1242-2	5.544	4.745	161.7E6	59212746	499.677	512.049
18)	L4 AR-1242-3	5.606	4.921	104.8E6	32647360	501.419	514.673
19)	L4 AR-1242-4	5.703	5.004	80827975	32698937	505.157	510.527
20)	L4 AR-1242-5	6.436	5.526	77500840	39282831	495.270	506.721



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101524\
Data File : PO107213.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2024 03:12
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO101524AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 04:37:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101524.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 16 04:36:18 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.373 min
Delta R.T.: 0.000 min
Response: 462467958
Conc: 51.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO101524AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: 0.000 min
Response: 169014013
Conc: 52.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.058 min
Delta R.T.: 0.003 min
Response: 124432920
Conc: 50.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.637 min
Delta R.T.: 0.000 min
Response: 137612293
Conc: 50.30 ng/ml



#16 AR-1242-1


R.T.: 5.522 min
Delta R.T.: 0.000 min
Response: 112547896
Conc: 503.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO101524AR1242




#16 AR-1242-1

R.T.: 4.726 min
Delta R.T.: 0.000 min
Response: 42391283
Conc: 506.20 ng/ml




#17 AR-1242-2

R.T.: 5.544 min
Delta R.T.: 0.000 min
Response: 161655864
Conc: 499.68 ng/ml



#17 AR-1242-2


R.T.: 4.745 min
Delta R.T.: 0.000 min
Response: 59212746
Conc: 512.05 ng/ml



#18 AR-1242-3


R.T.: 5.606 min
Delta R.T.: 0.000 min
Response: 104784876
Conc: 501.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO101524AR1242




#18 AR-1242-3

R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 32647360
Conc: 514.67 ng/ml




#19 AR-1242-4

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 80827975
Conc: 505.16 ng/ml



#19 AR-1242-4


R.T.: 5.004 min
Delta R.T.: 0.000 min
Response: 32698937
Conc: 510.53 ng/ml



#20 AR-1242-5

R.T.: 6.436 min
Delta R.T.: 0.000 min
Response: 77500840
Conc: 495.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO101524AR1242



#20 AR-1242-5

R.T.: 5.526 min
Delta R.T.: 0.000 min
Response: 39282831
Conc: 506.72 ng/ml