

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101619\
 Data File : P0062049.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Oct 2019 14:05
 Operator : HP/AJ
 Sample : K5111-08
 Misc : AR1232-LOQ-50PPB-WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOQ-WATER-02-QT4-2019

Manual Integrations
 APPROVED

Ankita
 10/17/2019 9:14:48 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 14:23:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 12 06:25:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.324	3.484	1236702	862121	19.996	20.105
2)	SA Decachlor...	9.954	8.338	1115430	584327	19.935	19.915
Target Compounds							
11)	L3 AR-1232-1	4.694	3.845	117223	85022	53.671	54.870
12)	L3 AR-1232-2	5.219	4.546	57033	87752	47.803	50.828m
13)	L3 AR-1232-3	5.507	4.715	125786	40893	49.169	44.739
14)	L3 AR-1232-4	5.667	4.797	58391	29859	45.423	36.703
15)	L3 AR-1232-5	5.755	4.963	39491	45957	39.901	52.037 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101619\
Data File : PO062049.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Oct 2019 14:05
Operator : HP/AJ
Sample : K5111-08
Misc : AR1232-LOQ-50PPB-WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT4-2019

Manual Integrations
APPROVED

Ankita
10/17/2019 9:14:48 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 14:23:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

