

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101620\
 Data File : PO072421.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Oct 2020 11:08
 Operator : DD\AJ
 Sample : L4406-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 JC-02-101420-A

Manual Integrations
 APPROVED

Ankita
 10/19/2020 3:06:46 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 05:45:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 14 17:47:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.319	3.503	2033162	986959	23.971m	23.431
2)	SA Decachlor...	9.922	8.672	2332947	1032830	21.150	19.979
Target Compounds							
26)	L6 AR-1254-1	6.515	5.576	217496	65797	51.485	23.803 #
27)	L6 AR-1254-2	6.741	5.735	245738	58066	37.917m	23.883 #
28)	L6 AR-1254-3	7.116	6.144	229136	106258	33.713	27.613
29)	L6 AR-1254-4	7.410	6.385	188501	86931	28.439m	32.374
30)	L6 AR-1254-5	7.831	6.805	258117	164094	40.295m	44.777

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101620\
Data File : PO072421.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Oct 2020 11:08
Operator : DD\AJ
Sample : L4406-01
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
JC-02-101420-A


Manual Integrations
APPROVED

Ankita
10/19/2020 3:06:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 05:45:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 14 17:47:24 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.319 min
Delta R.T.: -0.002 min
Response: 2033162
Conc: 23.97 ng/ml m

Instrument :
ECD_O
ClientSampleId :
JC-02-101420-A


Manual Integrations
APPROVED

Ankita
10/19/2020 3:06:46 PM




#1 Tetrachloro-m-xylene

R.T.: 3.503 min
Delta R.T.: 0.000 min
Response: 986959
Conc: 23.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.922 min
Delta R.T.: -0.004 min
Response: 2332947
Conc: 21.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.672 min
Delta R.T.: -0.003 min
Response: 1032830
Conc: 19.98 ng/ml




#26 AR-1254-1

R.T.: 6.515 min
Delta R.T.: 0.000 min
Response: 217496
Conc: 51.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
JC-02-101420-A


Manual Integrations
APPROVED

Ankita
10/19/2020 3:06:46 PM




#26 AR-1254-1

R.T.: 5.576 min
Delta R.T.: -0.002 min
Response: 65797
Conc: 23.80 ng/ml




#27 AR-1254-2

R.T.: 6.741 min
Delta R.T.: -0.003 min
Response: 245738
Conc: 37.92 ng/ml m



#27 AR-1254-2

R.T.: 5.735 min
Delta R.T.: -0.003 min
Response: 58066
Conc: 23.88 ng/ml




#28 AR-1254-3

R.T.: 7.116 min
Delta R.T.: -0.003 min
Response: 229136
Conc: 33.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
JC-02-101420-A


Manual Integrations
APPROVED

Ankita
10/19/2020 3:06:46 PM




#28 AR-1254-3

R.T.: 6.144 min
Delta R.T.: -0.003 min
Response: 106258
Conc: 27.61 ng/ml




#29 AR-1254-4

R.T.: 7.410 min
Delta R.T.: -0.004 min
Response: 188501
Conc: 28.44 ng/ml m



#29 AR-1254-4

R.T.: 6.385 min
Delta R.T.: -0.003 min
Response: 86931
Conc: 32.37 ng/ml




#30 AR-1254-5

R.T.: 7.831 min
Delta R.T.: -0.005 min
Response: 258117
Conc: 40.29 ng/ml m

Instrument :
ECD_O
ClientSampleId :
JC-02-101420-A

Manual Integrations
APPROVED

Ankita
10/19/2020 3:06:46 PM



#30 AR-1254-5

R.T.: 6.805 min
Delta R.T.: -0.003 min
Response: 164094
Conc: 44.78 ng/ml