

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101620\
 Data File : PO072472.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Oct 2020 3:26
 Operator : DD\AJ
 Sample : L4442-18
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 C0AL9

Manual Integrations
 APPROVED

Ankita
 10/19/2020 3:08:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 06:33:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 14 17:47:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.321	3.505	1674917	830488	19.747	19.716
2) SA Decachlor...	9.924	8.671	1487760	789201	13.488	15.267
Target Compounds						
31) L7 AR-1260-1	7.279	6.280	235698	181956	48.366m	71.126 #
32) L7 AR-1260-2	7.545	6.481	507332	268141	66.121m	83.156 #
33) L7 AR-1260-3	7.901	6.627	354194	171185	55.267m	57.038
34) L7 AR-1260-4	8.129	7.102	370548	129160	60.885m	48.989
35) L7 AR-1260-5	8.444	7.355	738332	326818	50.824	50.023

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101620\
Data File : PO072472.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Oct 2020 3:26
Operator : DD\AJ
Sample : L4442-18
Misc :
ALS Vial : 46 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
C0AL9


Manual Integrations
APPROVED

Ankita
10/19/2020 3:08:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 06:33:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 14 17:47:24 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.321 min
Delta R.T.: 0.000 min
Response: 1674917
Conc: 19.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL9


Manual Integrations
APPROVED

Ankita
10/19/2020 3:08:15 PM




#1 Tetrachloro-m-xylene

R.T.: 3.505 min
Delta R.T.: 0.000 min
Response: 830488
Conc: 19.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.924 min
Delta R.T.: -0.002 min
Response: 1487760
Conc: 13.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.671 min
Delta R.T.: -0.004 min
Response: 789201
Conc: 15.27 ng/ml




#31 AR-1260-1

R.T.: 7.279 min
Delta R.T.: -0.001 min
Response: 235698
Conc: 48.37 ng/ml m

Instrument :
ECD_O
ClientSampleId :
C0AL9


Manual Integrations
APPROVED

Ankita
10/19/2020 3:08:15 PM




#31 AR-1260-1

R.T.: 6.280 min
Delta R.T.: -0.002 min
Response: 181956
Conc: 71.13 ng/ml




#32 AR-1260-2

R.T.: 7.545 min
Delta R.T.: -0.002 min
Response: 507332
Conc: 66.12 ng/ml m



#32 AR-1260-2

R.T.: 6.481 min
Delta R.T.: -0.002 min
Response: 268141
Conc: 83.16 ng/ml




#33 AR-1260-3

R.T.: 7.901 min
Delta R.T.: -0.003 min
Response: 354194
Conc: 55.27 ng/ml m

Instrument :
ECD_O
ClientSampleId :
C0AL9


Manual Integrations
APPROVED

Ankita
10/19/2020 3:08:15 PM




#33 AR-1260-3

R.T.: 6.627 min
Delta R.T.: -0.002 min
Response: 171185
Conc: 57.04 ng/ml




#34 AR-1260-4

R.T.: 8.129 min
Delta R.T.: -0.001 min
Response: 370548
Conc: 60.88 ng/ml m



#34 AR-1260-4

R.T.: 7.102 min
Delta R.T.: -0.002 min
Response: 129160
Conc: 48.99 ng/ml




#35 AR-1260-5

R.T.: 8.444 min
Delta R.T.: -0.001 min
Response: 738332
Conc: 50.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL9

Manual Integrations
APPROVED

Ankita
10/19/2020 3:08:15 PM



#35 AR-1260-5

R.T.: 7.355 min
Delta R.T.: -0.003 min
Response: 326818
Conc: 50.02 ng/ml