

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101625\
 Data File : PO114435.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Oct 2025 15:37
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 06:28:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 06:46:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.357	3.660	299.6E6	225.5E6	55.177	56.339
2)	SA Decachlor...	9.944	8.669	184.7E6	157.6E6	50.720	47.289
Target Compounds							
21)	L5 AR-1248-1	5.496	4.742	50648087	49274617	514.935	500.273
22)	L5 AR-1248-2	5.767	4.978	72408452	65253136	504.361	486.768
23)	L5 AR-1248-3	5.967	5.021	79428133	68176474	500.141	488.574
24)	L5 AR-1248-4	6.363	5.191	92413492	80717268	472.002	491.088
25)	L5 AR-1248-5	6.402	5.583	86777612	84531015	433.467	498.478



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101625\
Data File : PO114435.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2025 15:37
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 06:28:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 06:46:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.357 min
Delta R.T.: 0.002 min
Response: 299552329
Conc: 55.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.660 min
Delta R.T.: 0.000 min
Response: 225483877
Conc: 56.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.944 min
Delta R.T.: 0.009 min
Response: 184675570
Conc: 50.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.669 min
Delta R.T.: -0.002 min
Response: 157623910
Conc: 47.29 ng/ml



#21 AR-1248-1


R.T.: 5.496 min
Delta R.T.: 0.004 min
Response: 50648087
Conc: 514.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 49274617
Conc: 500.27 ng/ml




#22 AR-1248-2

R.T.: 5.767 min
Delta R.T.: 0.004 min
Response: 72408452
Conc: 504.36 ng/ml



#22 AR-1248-2


R.T.: 4.978 min
Delta R.T.: 0.000 min
Response: 65253136
Conc: 486.77 ng/ml



#23 AR-1248-3


R.T.: 5.967 min
Delta R.T.: 0.004 min
Response: 79428133
Conc: 500.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.021 min
Delta R.T.: 0.000 min
Response: 68176474
Conc: 488.57 ng/ml




#24 AR-1248-4

R.T.: 6.363 min
Delta R.T.: 0.004 min
Response: 92413492
Conc: 472.00 ng/ml



#24 AR-1248-4


R.T.: 5.191 min
Delta R.T.: 0.000 min
Response: 80717268
Conc: 491.09 ng/ml



#25 AR-1248-5

R.T.: 6.402 min
Delta R.T.: 0.005 min
Response: 86777612
Conc: 433.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.583 min
Delta R.T.: 0.000 min
Response: 84531015
Conc: 498.48 ng/ml