

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101724\
 Data File : P0107282.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2024 16:54
 Operator : YP/AJ
 Sample : P4412-18
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 KR-SS18

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/18/2024
 Supervised By :Ankita Jodhani 10/18/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 06:21:59 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 16 04:53:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.374	3.647	95185990	35779459	10.444	11.117
2) SA Decachlor...	10.055	8.638	73134115	52754891	29.810m	19.242 #
Target Compounds						
31) L7 AR-1260-1	7.125	6.207	52227494	27541050	201.971m	177.507
32) L7 AR-1260-2	7.380	6.394	62576274	36720445	237.734m	208.502
33) L7 AR-1260-3	7.743	6.547	61641086	29337944	342.936	174.893 #
34) L7 AR-1260-4	7.966	7.018	63716546	26367828	363.049	182.248 #
35) L7 AR-1260-5	8.279	7.258	95939591	87979626	337.348	267.058

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101724\
Data File : PO107282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2024 16:54
Operator : YP/AJ
Sample : P4412-18
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
KR-SS18


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/18/2024
Supervised By :Ankita Jodhani 10/18/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 18 06:21:59 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101524.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 16 04:53:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.374 min
Delta R.T.: 0.000 min
Response: 95185990
Conc: 10.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
KR-SS18


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/18/2024
Supervised By :Ankita Jodhani 10/18/2024




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: 0.002 min
Response: 35779459
Conc: 11.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.055 min
Delta R.T.: -0.003 min
Response: 73134115
Conc: 29.81 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.638 min
Delta R.T.: 0.000 min
Response: 52754891
Conc: 19.24 ng/ml




#31 AR-1260-1

R.T.: 7.125 min
Delta R.T.: 0.000 min
Response: 52227494
Conc: 201.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
KR-SS18


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/18/2024
Supervised By :Ankita Jodhani 10/18/2024




#31 AR-1260-1

R.T.: 6.207 min
Delta R.T.: 0.000 min
Response: 27541050
Conc: 177.51 ng/ml




#32 AR-1260-2

R.T.: 7.380 min
Delta R.T.: -0.002 min
Response: 62576274
Conc: 237.73 ng/ml m



#32 AR-1260-2

R.T.: 6.394 min
Delta R.T.: 0.000 min
Response: 36720445
Conc: 208.50 ng/ml




#33 AR-1260-3

R.T.: 7.743 min
Delta R.T.: 0.000 min
Response: 61641086
Conc: 342.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
KR-SS18


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/18/2024
Supervised By :Ankita Jodhani 10/18/2024




#33 AR-1260-3

R.T.: 6.547 min
Delta R.T.: 0.000 min
Response: 29337944
Conc: 174.89 ng/ml




#34 AR-1260-4

R.T.: 7.966 min
Delta R.T.: -0.002 min
Response: 63716546
Conc: 363.05 ng/ml



#34 AR-1260-4

R.T.: 7.018 min
Delta R.T.: 0.000 min
Response: 26367828
Conc: 182.25 ng/ml




#35 AR-1260-5

R.T.: 8.279 min
Delta R.T.: -0.002 min
Response: 95939591
Conc: 337.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
KR-SS18

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/18/2024
Supervised By :Ankita Jodhani 10/18/2024



#35 AR-1260-5

R.T.: 7.258 min
Delta R.T.: -0.001 min
Response: 87979626
Conc: 267.06 ng/ml