

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0101819\
 Data File : P0062196.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Oct 2019 5:17
 Operator : HP/AJ
 Sample : K5436-01
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 3425-28

Manual Integrations
 APPROVED

Ankita
 10/21/2019 5:04:02 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 19 06:41:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 12 06:25:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.320	3.481	1343474	1149386	21.722m	26.804
2)	SA Decachlor...	9.948	8.332	1120528	564926	20.026	19.254
Target Compounds							
21)	L5 AR-1248-1	5.486	4.526	97160	56587	67.790m	57.639m
22)	L5 AR-1248-2	5.752	4.754	879003	682649	439.030m	538.834
23)	L5 AR-1248-3	5.954	4.793	1321798	648643	598.483m	488.637
24)	L5 AR-1248-4	6.356	4.959	1059846	921889	414.867m	582.599 #
25)	L5 AR-1248-5	6.393	5.340	2631148	1495635	1076.072m	971.822m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101819\
Data File : PO062196.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Oct 2019 5:17
Operator : HP/AJ
Sample : K5436-01
Misc :
ALS Vial : 49 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
3425-28


Manual Integrations
APPROVED

Ankita
10/21/2019 5:04:02 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 06:41:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.320 min
Delta R.T.: -0.005 min
Response: 1343474
Conc: 21.72 ng/ml m

Instrument :
ECD_O
ClientSampleId :
3425-28


Manual Integrations
APPROVED

Ankita
10/21/2019 5:04:02 PM




#1 Tetrachloro-m-xylene

R.T.: 3.481 min
Delta R.T.: 0.000 min
Response: 1149386
Conc: 26.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.948 min
Delta R.T.: -0.008 min
Response: 1120528
Conc: 20.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.332 min
Delta R.T.: -0.006 min
Response: 564926
Conc: 19.25 ng/ml




#21 AR-1248-1

R.T.: 5.486 min
Delta R.T.: -0.001 min
Response: 97160
Conc: 67.79 ng/ml m

Instrument :
ECD_O
ClientSampleId :
3425-28


Manual Integrations
APPROVED

Ankita
10/21/2019 5:04:02 PM




#21 AR-1248-1

R.T.: 4.526 min
Delta R.T.: -0.001 min
Response: 56587
Conc: 57.64 ng/ml m




#22 AR-1248-2

R.T.: 5.752 min
Delta R.T.: -0.006 min
Response: 879003
Conc: 439.03 ng/ml m



#22 AR-1248-2

R.T.: 4.754 min
Delta R.T.: -0.003 min
Response: 682649
Conc: 538.83 ng/ml




#23 AR-1248-3

R.T.: 5.954 min
Delta R.T.: -0.006 min
Response: 1321798
Conc: 598.48 ng/ml m

Instrument :
ECD_O
ClientSampleId :
3425-28


Manual Integrations
APPROVED

Ankita
10/21/2019 5:04:02 PM




#23 AR-1248-3

R.T.: 4.793 min
Delta R.T.: -0.004 min
Response: 648643
Conc: 488.64 ng/ml




#24 AR-1248-4

R.T.: 6.356 min
Delta R.T.: -0.007 min
Response: 1059846
Conc: 414.87 ng/ml m



#24 AR-1248-4

R.T.: 4.959 min
Delta R.T.: -0.003 min
Response: 921889
Conc: 582.60 ng/ml




#25 AR-1248-5

R.T.: 6.393 min
Delta R.T.: -0.008 min
Response: 2631148
Conc: 1076.07 ng/ml m

Instrument :
ECD_O
ClientSampleId :
3425-28

Manual Integrations
APPROVED

Ankita
10/21/2019 5:04:02 PM



#25 AR-1248-5

R.T.: 5.340 min
Delta R.T.: -0.004 min
Response: 1495635
Conc: 971.82 ng/ml m