

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0101821.M
 Title : GC EXTRACTABLES
 Last Update : Tue Oct 19 03:50:52 2021
 Response Via : Initial Calibration

Calibration Files

500 =P0082085.D 1000 =P0082083.D 750 =P0082084.D
 250 =P0082086.D 50 =P0082087.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	8.463	8.091	8.341	8.336	7.959	8.238	E4	2.51
2) SA	Decachlorobiphenyl	5.747	5.501	5.680	5.678	5.570	5.635	E4	1.75
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	1.543	1.413	1.475	1.525	1.483	1.488	E3	3.39
22) L5	AR-1248-2	1.956	1.804	1.883	1.953	2.126	1.945	E3	6.12
23) L5	AR-1248-3	2.259	2.103	2.189	2.265	2.046	2.172	E3	4.43
24) L5	AR-1248-4	2.431	2.251	2.335	2.442	2.337	2.359	E3	3.34
25) L5	AR-1248-5	2.304	2.158	2.230	2.318	2.132	2.228	E3	3.74
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0082085.D 1000 =P0082083.D 750 =P0082084.D
 250 =P0082086.D 50 =P0082087.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.187	2.991	3.100	3.129	3.237	3.129	E4	2.99
2) SA	Decachlorobiphenyl	2.823	2.583	2.688	2.929	3.134	2.832	E4	7.56
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0101821.M
 Title : GC EXTRACTABLES
 Last Update : Tue Oct 19 03:50:52 2021
 Response Via : Initial Calibration

Calibration Files

500 =P0082085.D 1000 =P0082083.D 750 =P0082084.D
 250 =P0082086.D 50 =P0082087.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	7.470	6.671	7.057	7.940	7.387	7.305 E2	6.50
22) L5	AR-1248-2	0.985	0.877	0.930	1.052	0.965	0.962 E3	6.76
23) L5	AR-1248-3	1.010	0.868	0.927	1.078	1.045	0.986 E3	8.80
24) L5	AR-1248-4	1.168	1.046	1.104	1.237	1.192	1.149 E3	6.53
25) L5	AR-1248-5	1.140	1.028	1.077	1.115	1.072	1.086 E3	3.95
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range