

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101918\
 Data File : P0050353.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Oct 2018 20:44
 Operator : SM/SJ
 Sample : J5577-04
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SSSI-1016-S-DH-21

Manual Integrations
 APPROVED

sohil
 10/30/2018 8:47:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 02:11:30 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 19 02:02:26 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.371	3.361	4987280	3547686	16.714	18.045
2) SA Decachlor...	10.114	8.116	3335577	3441005	15.372m	18.002

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO101918\
Data File : PO050353.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Oct 2018 20:44
Operator : SM/SJ
Sample : J5577-04
Misc :
ALS Vial : 44 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
SSSI-1016-S-DH-21


Manual Integrations
APPROVED

sohil
10/30/2018 8:47:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 02:11:30 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101818.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 19 02:02:26 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.371 min
Delta R.T.: 0.022 min
Response: 4987280
Conc: 16.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
SSSI-1016-S-DH-21


Manual Integrations
APPROVED

sohil
10/30/2018 8:47:22 AM




#1 Tetrachloro-m-xylene

R.T.: 3.361 min
Delta R.T.: 0.001 min
Response: 3547686
Conc: 18.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.114 min
Delta R.T.: 0.033 min
Response: 3335577
Conc: 15.37 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.116 min
Delta R.T.: 0.000 min
Response: 3441005
Conc: 18.00 ng/ml