

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102020\
 Data File : PO072584.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Oct 2020 5:23
 Operator : DD\AJ
 Sample : L4222-01MS
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 06:27:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 14 17:47:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.319	3.502	1968723	963783	23.211	22.881
2)	SA Decachlor...	9.919	8.667	2257104	1019390	20.463	19.719
Target Compounds							
20)	L4 AR-1242-5	0.000	5.572	0	35957	N.D.	26.190 #
24)	L5 AR-1248-4	6.527f	0.000	151666	0	34.297	N.D. #
26)	L6 AR-1254-1	6.527	5.572	151666	35957	35.902	13.008 #
27)	L6 AR-1254-2	6.746	0.000	227425	0	35.092	N.D. #
28)	L6 AR-1254-3	7.115	6.141	228231	60148	33.579	15.631 #
29)	L6 AR-1254-4	7.409	0.000	98131	0	14.805	N.D. #
30)	L6 AR-1254-5	7.829	6.803	352243	309559	54.989	84.470 #
31)	L7 AR-1260-1	7.276	6.277	270852	78989	55.580	30.876 #
32)	L7 AR-1260-2	7.553	6.478	1416603	130970	184.628	40.617 #
33)	L7 AR-1260-3	7.899	6.620	157680	72027	24.604	23.999
34)	L7 AR-1260-4	8.124	7.097	181550	44990	29.830	17.064 #
35)	L7 AR-1260-5	8.441	7.351	274063	158616	18.865	24.278 #
36)	L8 AR-1262-1	7.899	6.803	157680	309559	16.626	162.234 #
37)	L8 AR-1262-2	8.441	7.351	274063	158616	17.147	23.748 #
38)	L8 AR-1262-3	8.709	0.000	66246	0	9.480	N.D. #
39)	L8 AR-1262-4	8.781	7.690	258045	88722	64.571	17.215 #
40)	L8 AR-1262-5	9.336	0.000	163935	0	30.815	N.D. #
41)	L9 AR-1268-1	8.709	0.000	66246	0	3.441	N.D. #
42)	L9 AR-1268-2	8.781	7.690	258045	88722	15.630	12.194
44)	L9 AR-1268-4	9.336	0.000	163935	0	27.475	N.D. #
45)	L9 AR-1268-5	9.664	0.000	73242	0	1.781	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102020\
Data File : PO072584.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Oct 2020 5:23
Operator : DD\AJ
Sample : L4222-01MS
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 06:27:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 14 17:47:24 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.319 min
Delta R.T.: -0.002 min
Response: 1968723
Conc: 23.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.502 min
Delta R.T.: -0.002 min
Response: 963783
Conc: 22.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.919 min
Delta R.T.: -0.007 min
Response: 2257104
Conc: 20.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.667 min
Delta R.T.: -0.008 min
Response: 1019390
Conc: 19.72 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 6.585 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: -0.006 min
Response: 35957
Conc: 26.19 ng/ml




#24 AR-1248-4

R.T.: 6.527 min
Delta R.T.: -0.022 min
Response: 151666
Conc: 34.30 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 5.205 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.527 min
Delta R.T.: 0.013 min
Response: 151666
Conc: 35.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 5.572 min
Delta R.T.: -0.006 min
Response: 35957
Conc: 13.01 ng/ml




#27 AR-1254-2

R.T.: 6.746 min
Delta R.T.: 0.003 min
Response: 227425
Conc: 35.09 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T. : 5.738 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.115 min
Delta R.T.: -0.004 min
Response: 228231
Conc: 33.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 6.141 min
Delta R.T.: -0.006 min
Response: 60148
Conc: 15.63 ng/ml




#29 AR-1254-4

R.T.: 7.409 min
Delta R.T.: -0.005 min
Response: 98131
Conc: 14.80 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.388 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 7.829 min
Delta R.T.: -0.006 min
Response: 352243
Conc: 54.99 ng/ml

Instrument :
ECD_O
ClientSampleId :




#30 AR-1254-5

R.T.: 6.803 min
Delta R.T.: -0.005 min
Response: 309559
Conc: 84.47 ng/ml




#31 AR-1260-1

R.T.: 7.276 min
Delta R.T.: -0.005 min
Response: 270852
Conc: 55.58 ng/ml



#31 AR-1260-1


R.T.: 6.277 min
Delta R.T.: -0.005 min
Response: 78989
Conc: 30.88 ng/ml



#32 AR-1260-2


R.T.: 7.553 min
Delta R.T.: 0.006 min
Response: 1416603
Conc: 184.63 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.478 min
Delta R.T.: -0.006 min
Response: 130970
Conc: 40.62 ng/ml




#33 AR-1260-3

R.T.: 7.899 min
Delta R.T.: -0.005 min
Response: 157680
Conc: 24.60 ng/ml



#33 AR-1260-3


R.T.: 6.620 min
Delta R.T.: -0.009 min
Response: 72027
Conc: 24.00 ng/ml



#34 AR-1260-4


R.T.: 8.124 min
Delta R.T.: -0.006 min
Response: 181550
Conc: 29.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#34 AR-1260-4

R.T.: 7.097 min
Delta R.T.: -0.007 min
Response: 44990
Conc: 17.06 ng/ml




#35 AR-1260-5

R.T.: 8.441 min
Delta R.T.: -0.004 min
Response: 274063
Conc: 18.87 ng/ml



#35 AR-1260-5


R.T.: 7.351 min
Delta R.T.: -0.007 min
Response: 158616
Conc: 24.28 ng/ml



#36 AR-1262-1


R.T.: 7.899 min
Delta R.T.: -0.004 min
Response: 157680
Conc: 16.63 ng/ml

Instrument :
ECD_O
ClientSampleId :




#36 AR-1262-1

R.T.: 6.803 min
Delta R.T.: -0.003 min
Response: 309559
Conc: 162.23 ng/ml




#37 AR-1262-2

R.T.: 8.441 min
Delta R.T.: -0.004 min
Response: 274063
Conc: 17.15 ng/ml



#37 AR-1262-2


R.T.: 7.351 min
Delta R.T.: -0.006 min
Response: 158616
Conc: 23.75 ng/ml



#38 AR-1262-3


R.T.: 8.709 min
Delta R.T.: -0.004 min
Response: 66246
Conc: 9.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 7.635 min
Response: 0
Conc: N.D.




#39 AR-1262-4

R.T.: 8.781 min
Delta R.T.: -0.011 min
Response: 258045
Conc: 64.57 ng/ml



#39 AR-1262-4


R.T.: 7.690 min
Delta R.T.: -0.008 min
Response: 88722
Conc: 17.21 ng/ml



#40 AR-1262-5


R.T.: 9.336 min
Delta R.T.: -0.006 min
Response: 163935
Conc: 30.82 ng/ml

Instrument :
ECD_O
ClientSampleId :




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 8.195 min
Response: 0
Conc: N.D.




#41 AR-1268-1

R.T.: 8.709 min
Delta R.T.: -0.004 min
Response: 66246
Conc: 3.44 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T. : 7.635 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 8.781 min
Delta R.T.: -0.013 min
Response: 258045
Conc: 15.63 ng/ml

Instrument :
ECD_O
ClientSampleId :




#42 AR-1268-2

R.T.: 7.690 min
Delta R.T.: -0.012 min
Response: 88722
Conc: 12.19 ng/ml




#44 AR-1268-4

R.T.: 9.336 min
Delta R.T.: -0.005 min
Response: 163935
Conc: 27.48 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 8.195 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.664 min
Delta R.T.: -0.003 min
Response: 73242
Conc: 1.78 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.462 min
Response: 0
Conc: N.D.