

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102020\
 Data File : PO072527.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Oct 2020 11:12
 Operator : DD\AJ
 Sample : L4442-10DL2 200X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 C0AL1DL2

Manual Integrations
 APPROVED

Ankita
 10/21/2020 5:15:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 14:37:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 14 17:47:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
16)	L4	AR-1242-1	5.621	4.721	963144	376742 395.480 291.476m#
17)	L4	AR-1242-2	5.642	4.738	3132272	1463336 889.871 799.098m
18)	L4	AR-1242-3	5.706	4.922	461934	555694 223.292 590.279 #
19)	L4	AR-1242-4	5.812	5.020	1163638	623537 671.163 703.942
20)	L4	AR-1242-5	6.581	5.572	653759	855703 280.439m 623.271 #
31)	L7	AR-1260-1	7.275	6.277	1653661	764595 339.337 298.878
32)	L7	AR-1260-2	7.542	6.478	2419621	853027 315.353 264.542
33)	L7	AR-1260-3	7.898	6.622	949997	687560 148.235 229.090 #
34)	L7	AR-1260-4	8.124	7.098	1194689	319990 196.299 121.369 #
35)	L7	AR-1260-5	8.440	7.351	2025825	858987 139.450 131.478

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102020\
Data File : PO072527.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Oct 2020 11:12
Operator : DD\AJ
Sample : L4442-10DL2 200X
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
C0AL1DL2


Manual Integrations
APPROVED

Ankita
10/21/2020 5:15:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 14:37:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 14 17:47:24 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.321 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
C0AL1DL2


Manual Integrations
APPROVED

Ankita
10/21/2020 5:15:06 PM




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.504 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.926 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.675 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.621 min
Delta R.T.: -0.003 min
Response: 963144
Conc: 395.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL1DL2


Manual Integrations
APPROVED

Ankita
10/21/2020 5:15:06 PM




#16 AR-1242-1

R.T.: 4.721 min
Delta R.T.: -0.004 min
Response: 376742
Conc: 291.48 ng/ml m




#17 AR-1242-2

R.T.: 5.642 min
Delta R.T.: -0.004 min
Response: 3132272
Conc: 889.87 ng/ml



#17 AR-1242-2

R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 1463336
Conc: 799.10 ng/ml m




#18 AR-1242-3

R.T.: 5.706 min
Delta R.T.: -0.002 min
Response: 461934
Conc: 223.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL1DL2


Manual Integrations
APPROVED

Ankita
10/21/2020 5:15:06 PM




#18 AR-1242-3

R.T.: 4.922 min
Delta R.T.: -0.005 min
Response: 555694
Conc: 590.28 ng/ml




#19 AR-1242-4

R.T.: 5.812 min
Delta R.T.: -0.004 min
Response: 1163638
Conc: 671.16 ng/ml



#19 AR-1242-4

R.T.: 5.020 min
Delta R.T.: -0.005 min
Response: 623537
Conc: 703.94 ng/ml




#20 AR-1242-5

R.T.: 6.581 min
Delta R.T.: -0.004 min
Response: 653759
Conc: 280.44 ng/ml m

Instrument :
ECD_O
ClientSampleId :
C0AL1DL2


Manual Integrations
APPROVED

Ankita
10/21/2020 5:15:06 PM




#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: -0.006 min
Response: 855703
Conc: 623.27 ng/ml




#31 AR-1260-1

R.T.: 7.275 min
Delta R.T.: -0.005 min
Response: 1653661
Conc: 339.34 ng/ml



#31 AR-1260-1

R.T.: 6.277 min
Delta R.T.: -0.005 min
Response: 764595
Conc: 298.88 ng/ml




#32 AR-1260-2

R.T.: 7.542 min
Delta R.T.: -0.005 min
Response: 2419621
Conc: 315.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL1DL2


Manual Integrations
APPROVED

Ankita
10/21/2020 5:15:06 PM




#32 AR-1260-2

R.T.: 6.478 min
Delta R.T.: -0.006 min
Response: 853027
Conc: 264.54 ng/ml




#33 AR-1260-3

R.T.: 7.898 min
Delta R.T.: -0.006 min
Response: 949997
Conc: 148.23 ng/ml



#33 AR-1260-3

R.T.: 6.622 min
Delta R.T.: -0.006 min
Response: 687560
Conc: 229.09 ng/ml




#34 AR-1260-4

R.T.: 8.124 min
Delta R.T.: -0.006 min
Response: 1194689
Conc: 196.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL1DL2


Manual Integrations
APPROVED

Ankita
10/21/2020 5:15:06 PM




#34 AR-1260-4

R.T.: 7.098 min
Delta R.T.: -0.005 min
Response: 319990
Conc: 121.37 ng/ml



#35 AR-1260-5

R.T.: 8.440 min
Delta R.T.: -0.005 min
Response: 2025825
Conc: 139.45 ng/ml



#35 AR-1260-5

R.T.: 7.351 min
Delta R.T.: -0.007 min
Response: 858987
Conc: 131.48 ng/ml