

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102025\
 Data File : PO114514.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2025 23:28
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 00:11:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 18 06:07:58 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.355	3.657	287.3E6	205.7E6	52.917	51.389
2)	SA Decachlor...	9.942	8.661	178.9E6	152.6E6	49.141	45.774
Target Compounds							
3)	L1 AR-1016-1	5.494	4.738	93892142	85091618	520.588	505.341
4)	L1 AR-1016-2	5.515	4.758	150.1E6	122.2E6	493.387	505.090
5)	L1 AR-1016-3	5.577	4.932	87217865	66159959	493.168	532.888
6)	L1 AR-1016-4	5.673	4.974	72370234	51100879	526.120	529.403
7)	L1 AR-1016-5	5.964	5.188	64583689	67559184	511.491	528.403
31)	L7 AR-1260-1	7.078	6.220	111.2E6	132.0E6	510.454	560.662
32)	L7 AR-1260-2	7.331	6.407	152.7E6	188.3E6	494.921	563.001
33)	L7 AR-1260-3	7.690	6.560	117.5E6	150.8E6	458.283	511.148
34)	L7 AR-1260-4	7.911	7.032	115.1E6	126.5E6	493.971	567.489
35)	L7 AR-1260-5	8.217	7.272	281.8E6	326.6E6	485.806	532.854



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102025\
Data File : PO114514.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2025 23:28
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 00:11:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 18 06:07:58 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.355 min
Delta R.T.: 0.000 min
Response: 287282419
Conc: 52.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.657 min
Delta R.T.: -0.002 min
Response: 205671908
Conc: 51.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.942 min
Delta R.T.: 0.007 min
Response: 178926565
Conc: 49.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.661 min
Delta R.T.: -0.010 min
Response: 152573462
Conc: 45.77 ng/ml



#3 AR-1016-1


R.T.: 5.494 min
Delta R.T.: 0.002 min
Response: 93892142
Conc: 520.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 85091618
Conc: 505.34 ng/ml




#4 AR-1016-2

R.T.: 5.515 min
Delta R.T.: 0.002 min
Response: 150136411
Conc: 493.39 ng/ml



#4 AR-1016-2


R.T.: 4.758 min
Delta R.T.: -0.004 min
Response: 122249817
Conc: 505.09 ng/ml



#5 AR-1016-3


R.T.: 5.577 min
Delta R.T.: 0.001 min
Response: 87217865
Conc: 493.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.932 min
Delta R.T.: -0.004 min
Response: 66159959
Conc: 532.89 ng/ml




#6 AR-1016-4

R.T.: 5.673 min
Delta R.T.: 0.001 min
Response: 72370234
Conc: 526.12 ng/ml



#6 AR-1016-4


R.T.: 4.974 min
Delta R.T.: -0.005 min
Response: 51100879
Conc: 529.40 ng/ml



#7 AR-1016-5


R.T.: 5.964 min
Delta R.T.: 0.001 min
Response: 64583689
Conc: 511.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.188 min
Delta R.T.: -0.004 min
Response: 67559184
Conc: 528.40 ng/ml




#31 AR-1260-1

R.T.: 7.078 min
Delta R.T.: 0.003 min
Response: 111243878
Conc: 510.45 ng/ml



#31 AR-1260-1


R.T.: 6.220 min
Delta R.T.: -0.005 min
Response: 132026538
Conc: 560.66 ng/ml



#32 AR-1260-2


R.T.: 7.331 min
Delta R.T.: 0.003 min
Response: 152685421
Conc: 494.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.407 min
Delta R.T.: -0.006 min
Response: 188261079
Conc: 563.00 ng/ml




#33 AR-1260-3

R.T.: 7.690 min
Delta R.T.: 0.005 min
Response: 117535939
Conc: 458.28 ng/ml



#33 AR-1260-3


R.T.: 6.560 min
Delta R.T.: -0.006 min
Response: 150758700
Conc: 511.15 ng/ml



#34 AR-1260-4


R.T.: 7.911 min
Delta R.T.: 0.004 min
Response: 115055548
Conc: 493.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.032 min
Delta R.T.: -0.007 min
Response: 126498579
Conc: 567.49 ng/ml



#35 AR-1260-5

R.T.: 8.217 min
Delta R.T.: 0.005 min
Response: 281845198
Conc: 485.81 ng/ml



#35 AR-1260-5

R.T.: 7.272 min
Delta R.T.: -0.007 min
Response: 326618997
Conc: 532.85 ng/ml