

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102119\
 Data File : PO063033.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Oct 2019 15:29
 Operator : HP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:16:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 12 06:25:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.316	3.476	1244094	913270	20.116	21.297
2)	SA Decachlor...	9.940	8.327	1200633	654708	21.458	22.313
Target Compounds							
27)	L6 AR-1254-2	6.591	5.513	67479	63470	16.505	29.969 #
30)	L6 AR-1254-5	7.593	0.000	537716	0	171.581	N.D. #
34)	L7 AR-1260-4	7.924	0.000	48861	0	17.469	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102119\
Data File : P0063033.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Oct 2019 15:29
Operator : HP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:16:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 12 06:25:38 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.316 min
Delta R.T.: -0.009 min
Response: 1244094
Conc: 20.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.476 min
Delta R.T.: -0.005 min
Response: 913270
Conc: 21.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.940 min
Delta R.T.: -0.016 min
Response: 1200633
Conc: 21.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.327 min
Delta R.T.: -0.011 min
Response: 654708
Conc: 22.31 ng/ml



#27 AR-1254-2


R.T.: 6.591 min
Delta R.T.: 0.038 min
Response: 67479
Conc: 16.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.513 min
Delta R.T.: 0.064 min
Response: 63470
Conc: 29.97 ng/ml




#30 AR-1254-5

R.T.: 7.593 min
Delta R.T.: -0.027 min
Response: 537716
Conc: 171.58 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 6.476 min
Response: 0
Conc: N.D.



#34 AR-1260-4

R.T.: 7.924 min
Delta R.T.: 0.005 min
Response: 48861
Conc: 17.47 ng/ml

Instrument :
ECD_O
ClientSampleId :



#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 6.766 min
Response: 0
Conc: N.D.