

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102220\
 Data File : PO072655.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Oct 2020 12:51
 Operator : DD\AJ
 Sample : PB132443BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 05:50:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 14 17:47:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.317	3.500	2224506	1094305	26.227	25.979
2)	SA Decachlor...	9.916	8.663	2551373	1213981	23.130	23.484
Target Compounds							
24)	L5 AR-1248-4	6.521f	0.000	91870	0	20.775	N.D. #
26)	L6 AR-1254-1	6.521	0.000	91870	0	21.747	N.D. #
28)	L6 AR-1254-3	7.144f	0.000	80031	0	11.775	N.D. #
29)	L6 AR-1254-4	7.412	0.000	23865	0	3.601	N.D. #
35)	L7 AR-1260-5	8.442	0.000	42736	0	2.942	N.D. #
37)	L8 AR-1262-2	8.442	0.000	42736	0	2.674	N.D. #
39)	L8 AR-1262-4	8.782	0.000	69668	0	17.433	N.D. #
42)	L9 AR-1268-2	8.782	0.000	69668	0	4.220	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102220\
Data File : PO072655.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Oct 2020 12:51
Operator : DD\AJ
Sample : PB132443BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 05:50:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 14 17:47:24 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.317 min
Delta R.T.: -0.004 min
Response: 2224506
Conc: 26.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.500 min
Delta R.T.: -0.004 min
Response: 1094305
Conc: 25.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.916 min
Delta R.T.: -0.010 min
Response: 2551373
Conc: 23.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.663 min
Delta R.T.: -0.012 min
Response: 1213981
Conc: 23.48 ng/ml



#24 AR-1248-4


R.T.: 6.521 min
Delta R.T.: -0.029 min
Response: 91870
Conc: 20.77 ng/ml

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.205 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.521 min
Delta R.T.: 0.006 min
Response: 91870
Conc: 21.75 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.578 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.144 min
Delta R.T.: 0.025 min
Response: 80031
Conc: 11.77 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.146 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 7.412 min
Delta R.T.: -0.002 min
Response: 23865
Conc: 3.60 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 6.388 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.442 min
Delta R.T.: -0.004 min
Response: 42736
Conc: 2.94 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.358 min
Response: 0
Conc: N.D.




#37 AR-1262-2

R.T.: 8.442 min
Delta R.T.: -0.004 min
Response: 42736
Conc: 2.67 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 7.357 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.782 min
Delta R.T.: -0.011 min
Response: 69668
Conc: 17.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.698 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 8.782 min
Delta R.T.: -0.012 min
Response: 69668
Conc: 4.22 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 7.701 min
Response: 0
Conc: N.D.