

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0102220\
 Data File : P0072662.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Oct 2020 16:55
 Operator : DD\AJ
 Sample : L4214-01
 Misc : AR1262 LOD 25 PPB
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOD-MDL-SOIL-01-QT4-2020

Manual Integrations
 APPROVED

mohammad
 10/26/2020 5:11:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 05:53:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 14 17:47:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.316	3.499	1374879	704223	16.210	16.719
2) SA Decachlor...	9.918	8.664	1708113	918717	15.485	17.772
Target Compounds						
36) L8 AR-1262-1	7.898	6.798	270816	113842	28.556	59.662 #
37) L8 AR-1262-2	8.439	7.348	445037	236943	27.844	35.475 #
38) L8 AR-1262-3	8.707	7.627	150751	80274	21.573	28.167 #
39) L8 AR-1262-4	8.787	7.689	78687	174648	19.690m	33.887 #
40) L8 AR-1262-5	9.335	8.186	137755	71348	25.894	28.251

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102220\
Data File : PO072662.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Oct 2020 16:55
Operator : DD\AJ
Sample : L4214-01
Misc : AR1262 LOD 25 PPB
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleID :
LOD-MDL-SOIL-01-QT4-2020

Manual Integrations
APPROVED

mohammad
10/26/2020 5:11:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 05:53:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101420.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 14 17:47:24 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

