

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0102523\
 Data File : P0098938.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2023 21:18
 Operator : YP/AJ
 Sample : 04991-03
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 1025

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/26/2023
 Supervised By :Ankita Jodhani 10/26/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 26 00:47:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.475	3.631	40470225	15466533	20.482m	21.491
2) SA Decachlor...	10.280	8.630	22175688	10668904	19.575	19.885

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102523\
Data File : PO098938.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2023 21:18
Operator : YP/AJ
Sample : 04991-03
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
1025


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/26/2023
Supervised By :Ankita Jodhani 10/26/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 26 00:47:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.475 min
Delta R.T.: -0.003 min
Response: 40470225
Conc: 20.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
1025


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/26/2023
Supervised By :Ankita Jodhani 10/26/2023




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 15466533
Conc: 21.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.280 min
Delta R.T.: 0.000 min
Response: 22175688
Conc: 19.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.630 min
Delta R.T.: 0.000 min
Response: 10668904
Conc: 19.88 ng/ml