

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0102524\
 Data File : P0107383.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2024 16:07
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 23:59:22 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 16 04:53:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.373	3.645	498.6E6	182.8E6	54.707	56.797
2) SA Decachlor...	10.061	8.637	130.0E6	149.8E6	52.996	54.626
Target Compounds						
16) L4 AR-1242-1	5.522	4.725	120.5E6	47176098	538.709	563.341
17) L4 AR-1242-2	5.544	4.744	172.4E6	66114693	532.836	571.734
18) L4 AR-1242-3	5.607	4.920	107.4E6	34739669	514.113	547.657
19) L4 AR-1242-4	5.703	5.003	85445469	33674696	534.016	525.762
20) L4 AR-1242-5	6.437	5.525	79755281	39888384	509.677	514.533



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102524\
Data File : PO107383.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2024 16:07
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 23:59:22 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101524.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 16 04:53:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.373 min
Delta R.T.: -0.001 min
Response: 498590183
Conc: 54.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.645 min
Delta R.T.: 0.000 min
Response: 182800264
Conc: 56.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.061 min
Delta R.T.: 0.002 min
Response: 130019453
Conc: 53.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.637 min
Delta R.T.: -0.001 min
Response: 149763402
Conc: 54.63 ng/ml



#16 AR-1242-1


R.T.: 5.522 min
Delta R.T.: 0.000 min
Response: 120477678
Conc: 538.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.725 min
Delta R.T.: 0.000 min
Response: 47176098
Conc: 563.34 ng/ml




#17 AR-1242-2

R.T.: 5.544 min
Delta R.T.: -0.001 min
Response: 172383495
Conc: 532.84 ng/ml



#17 AR-1242-2


R.T.: 4.744 min
Delta R.T.: 0.000 min
Response: 66114693
Conc: 571.73 ng/ml



#18 AR-1242-3


R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 107437653
Conc: 514.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.920 min
Delta R.T.: 0.000 min
Response: 34739669
Conc: 547.66 ng/ml




#19 AR-1242-4

R.T.: 5.703 min
Delta R.T.: -0.001 min
Response: 85445469
Conc: 534.02 ng/ml



#19 AR-1242-4


R.T.: 5.003 min
Delta R.T.: 0.000 min
Response: 33674696
Conc: 525.76 ng/ml



#20 AR-1242-5

R.T.: 6.437 min
Delta R.T.: 0.000 min
Response: 79755281
Conc: 509.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.525 min
Delta R.T.: -0.001 min
Response: 39888384
Conc: 514.53 ng/ml