

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO102620\
 Data File : PO072789.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Oct 2020 23:45
 Operator : DD\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 ICVPO102620AR1242

Manual Integrations
 APPROVED

Ankita
 10/27/2020 10:00:27 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 05:37:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO102620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 27 05:29:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.948	4.020	2134433	1664784	46.293	47.501
2) SA Decachlor...	10.985	9.311	1941273	1840432	52.545m	55.187
Target Compounds						
16) L4 AR-1242-1	6.272	5.274	637062	543298	457.914	462.079
17) L4 AR-1242-2	6.295	5.295	881463	724018	461.100	455.914
18) L4 AR-1242-3	6.362	5.485	545876	401984	460.982	465.010
19) L4 AR-1242-4	6.469	5.582	452856	400136	467.757	482.091
20) L4 AR-1242-5	7.255	6.143	482408	517331	470.939	482.095

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102620\
Data File : PO072789.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Oct 2020 23:45
Operator : DD\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO102620AR1242

Manual Integrations
APPROVED

Ankita
10/27/2020 10:00:27 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:37:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102620.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 27 05:29:40 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

