

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0102621\
 Data File : P0082198.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Oct 2021 22:46
 Operator : AJ\MA
 Sample : M4358-07
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TH-GS

Manual Integrations
 APPROVED

mohammad
 10/27/2021 2:35:41 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 02:35:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0101821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 19 04:37:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.979	3.972	1984317	725844	24.558	24.601
2) SA Decachlor...	11.102	9.301	1069868	484835	19.249	19.036
Target Compounds						
31) L7 AR-1260-1	8.014	6.830	67361	24706	23.867m	14.925m#
32) L7 AR-1260-2	8.282	7.029	113627	60779	34.156	30.301
33) L7 AR-1260-3	8.652	7.183	56809	45637	25.204	24.090
34) L7 AR-1260-4	8.883	7.668	64932	25148	24.527m	18.347 #
35) L7 AR-1260-5	9.219	7.917	116834	77535	21.326	23.466

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102621\
Data File : P0082198.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Oct 2021 22:46
Operator : AJ\MA
Sample : M4358-07
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
TH-GS


Manual Integrations
APPROVED

mohammad
10/27/2021 2:35:41 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 02:35:15 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101821.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 19 04:37:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.979 min
Delta R.T.: -0.004 min
Response: 1984317
Conc: 24.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
TH-GS


Manual Integrations
APPROVED

mohammad
10/27/2021 2:35:41 PM




#1 Tetrachloro-m-xylene

R.T.: 3.972 min
Delta R.T.: -0.004 min
Response: 725844
Conc: 24.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.102 min
Delta R.T.: 0.000 min
Response: 1069868
Conc: 19.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.301 min
Delta R.T.: -0.006 min
Response: 484835
Conc: 19.04 ng/ml




#31 AR-1260-1

R.T.: 8.014 min
Delta R.T.: -0.003 min
Response: 67361
Conc: 23.87 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TH-GS


Manual Integrations
APPROVED

mohammad
10/27/2021 2:35:41 PM




#31 AR-1260-1

R.T.: 6.830 min
Delta R.T.: -0.005 min
Response: 24706
Conc: 14.93 ng/ml m




#32 AR-1260-2

R.T.: 8.282 min
Delta R.T.: -0.002 min
Response: 113627
Conc: 34.16 ng/ml



#32 AR-1260-2

R.T.: 7.029 min
Delta R.T.: -0.005 min
Response: 60779
Conc: 30.30 ng/ml




#33 AR-1260-3

R.T.: 8.652 min
Delta R.T.: 0.000 min
Response: 56809
Conc: 25.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
TH-GS


Manual Integrations
APPROVED

mohammad
10/27/2021 2:35:41 PM




#33 AR-1260-3

R.T.: 7.183 min
Delta R.T.: -0.006 min
Response: 45637
Conc: 24.09 ng/ml




#34 AR-1260-4

R.T.: 8.883 min
Delta R.T.: -0.002 min
Response: 64932
Conc: 24.53 ng/ml m



#34 AR-1260-4

R.T.: 7.668 min
Delta R.T.: -0.007 min
Response: 25148
Conc: 18.35 ng/ml




#35 AR-1260-5

R.T.: 9.219 min
Delta R.T.: -0.002 min
Response: 116834
Conc: 21.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
TH-GS

Manual Integrations
APPROVED

mohammad
10/27/2021 2:35:41 PM



#35 AR-1260-5

R.T.: 7.917 min
Delta R.T.: -0.007 min
Response: 77535
Conc: 23.47 ng/ml