

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO102623\
 Data File : P0099016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2023 06:14
 Operator : YP/AJ
 Sample : 05075-09
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 RWG-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/27/2023
 Supervised By :Ankita Jodhani 10/27/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 06:56:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.475	3.628	42292571	15795316	21.404	21.948
2) SA Decachlor...	10.276	8.627	26676843	12830664	23.548	23.914
Target Compounds						
21) L5 AR-1248-1	5.648	4.710	9929628	3715296	292.283m	268.059
22) L5 AR-1248-2	5.923	4.945	5551328	3267567	106.076m	164.500m#
23) L5 AR-1248-3	6.127	4.987	10315570	2818585	181.046m	135.161m#
24) L5 AR-1248-4	6.531	5.159	16577780	4485595	304.003m	182.464 #
25) L5 AR-1248-5	6.571	5.552	17316793	6354188	304.840m	292.572

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102623\
Data File : PO099016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2023 06:14
Operator : YP/AJ
Sample : 05075-09
Misc :
ALS Vial : 48 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
RWG-1

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 10/27/2023
Supervised By : Ankita Jodhani 10/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 06:56:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

