

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102722\
 Data File : PO089966.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2022 18:48
 Operator : YP/AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 01:27:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 01:25:03 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.306	3.486	322.5E6	110.4E6	97.072	96.320
2) SA Decachlor...	10.053	8.495	235.4E6	96600441	92.108	91.864
Target Compounds						
21) L5 AR-1248-1	5.481	4.573	60366423	21770867	921.515	898.609
22) L5 AR-1248-2	5.757	4.811	87587960	30454164	911.328	903.034
23) L5 AR-1248-3	5.961	4.852	93921089	31399628	921.076	902.371
24) L5 AR-1248-4	6.368	5.023	99783159	37055832	925.499	917.181
25) L5 AR-1248-5	6.406	5.418	97577620	34256746	926.944	919.041



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102722\
Data File : PO089966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2022 18:48
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:27:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 01:25:03 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.306 min
Delta R.T.: -0.001 min
Response: 322521308
Conc: 97.07 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.486 min
Delta R.T.: 0.000 min
Response: 110351343
Conc: 96.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.053 min
Delta R.T.: 0.003 min
Response: 235365075
Conc: 92.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.495 min
Delta R.T.: 0.001 min
Response: 96600441
Conc: 91.86 ng/ml



#21 AR-1248-1


R.T.: 5.481 min
Delta R.T.: 0.000 min
Response: 60366423
Conc: 921.52 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.573 min
Delta R.T.: 0.000 min
Response: 21770867
Conc: 898.61 ng/ml




#22 AR-1248-2

R.T.: 5.757 min
Delta R.T.: 0.000 min
Response: 87587960
Conc: 911.33 ng/ml



#22 AR-1248-2


R.T.: 4.811 min
Delta R.T.: 0.000 min
Response: 30454164
Conc: 903.03 ng/ml



#23 AR-1248-3


R.T.: 5.961 min
Delta R.T.: 0.000 min
Response: 93921089
Conc: 921.08 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.852 min
Delta R.T.: 0.000 min
Response: 31399628
Conc: 902.37 ng/ml




#24 AR-1248-4

R.T.: 6.368 min
Delta R.T.: 0.000 min
Response: 99783159
Conc: 925.50 ng/ml



#24 AR-1248-4


R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 37055832
Conc: 917.18 ng/ml



#25 AR-1248-5

R.T.: 6.406 min
Delta R.T.: 0.000 min
Response: 97577620
Conc: 926.94 ng/ml

Instrument :
ECD_O
ClientSampleId :



#25 AR-1248-5

R.T.: 5.418 min
Delta R.T.: 0.000 min
Response: 34256746
Conc: 919.04 ng/ml