

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102722\
 Data File : PO089967.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2022 19:05
 Operator : YP/AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 01:27:51 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 01:25:03 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.305	3.485	250.1E6	85711772	75.267	74.813
2)	SA Decachlor...	10.051	8.494	186.1E6	76357726	72.816	72.614
Target Compounds							
21)	L5 AR-1248-1	5.480	4.572	47446903	16971373	724.294	700.506
22)	L5 AR-1248-2	5.756	4.810	65820473	24230167	684.843	718.478
23)	L5 AR-1248-3	5.961	4.852	74405829	24986984	729.691	718.083
24)	L5 AR-1248-4	6.367	5.023	78880300	29342383	731.623	726.263
25)	L5 AR-1248-5	6.405	5.418	77260526	27156041	733.941	728.543



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102722\
Data File : PO089967.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2022 19:05
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:27:51 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 01:25:03 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.305 min
Delta R.T.: -0.002 min
Response: 250073845
Conc: 75.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.485 min
Delta R.T.: 0.000 min
Response: 85711772
Conc: 74.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.051 min
Delta R.T.: 0.001 min
Response: 186065838
Conc: 72.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.494 min
Delta R.T.: 0.000 min
Response: 76357726
Conc: 72.61 ng/ml



#21 AR-1248-1


R.T.: 5.480 min
Delta R.T.: 0.000 min
Response: 47446903
Conc: 724.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 4.572 min
Delta R.T.: 0.000 min
Response: 16971373
Conc: 700.51 ng/ml




#22 AR-1248-2

R.T.: 5.756 min
Delta R.T.: -0.001 min
Response: 65820473
Conc: 684.84 ng/ml



#22 AR-1248-2


R.T.: 4.810 min
Delta R.T.: 0.000 min
Response: 24230167
Conc: 718.48 ng/ml



#23 AR-1248-3


R.T.: 5.961 min
Delta R.T.: 0.000 min
Response: 74405829
Conc: 729.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 4.852 min
Delta R.T.: 0.000 min
Response: 24986984
Conc: 718.08 ng/ml




#24 AR-1248-4

R.T.: 6.367 min
Delta R.T.: -0.001 min
Response: 78880300
Conc: 731.62 ng/ml



#24 AR-1248-4


R.T.: 5.023 min
Delta R.T.: 0.000 min
Response: 29342383
Conc: 726.26 ng/ml



#25 AR-1248-5

R.T.: 6.405 min
Delta R.T.: -0.001 min
Response: 77260526
Conc: 733.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.418 min
Delta R.T.: 0.000 min
Response: 27156041
Conc: 728.54 ng/ml