

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0102723\
 Data File : P0099041.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2023 17:46
 Operator : YP/AJ
 Sample : 05092-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 00:41:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.475	3.630	20495828	7677582	10.373	10.668
2)	SA Decachlor...	10.277	8.627	7884366	3635008	6.960	6.775
Target Compounds							
3)	L1 AR-1016-1	5.655	4.725	111028	119937	2.016	5.356 #
4)	L1 AR-1016-2	5.655	4.734	111028	58477	1.343	1.871 #
5)	L1 AR-1016-3	5.736	4.898	151393	111200	2.941	6.567 #
6)	L1 AR-1016-4	5.833	4.955	31410	98122	0.773	6.787 #
7)	L1 AR-1016-5	6.137	5.160	74580	47010	1.757	2.516 #
8)	L2 AR-1221-1	4.684	3.812	1097178	24090	44.191	2.593 #
9)	L2 AR-1221-2	4.784	3.910	660106	613635	35.683	94.642 #
10)	L2 AR-1221-3	4.849	4.005	69918	27631	1.291	1.366
11)	L3 AR-1232-1	4.849	4.005	69918	27631	1.528	1.612
12)	L3 AR-1232-2	5.404	4.734	83846	58477	3.551	3.817
13)	L3 AR-1232-3	5.655	4.898	111028	111200	2.811	13.897 #
14)	L3 AR-1232-4	5.833	5.003	31410	122074	1.646	15.867 #
15)	L3 AR-1232-5	5.915	5.160	24081	47010	1.446	5.527 #
16)	L4 AR-1242-1	5.655	4.725	111028	119937	2.565	6.826 #
17)	L4 AR-1242-2	5.655	4.734	111028	58477	1.734	2.413 #
18)	L4 AR-1242-3	5.736	4.898	151393	111200	3.764	8.673 #
19)	L4 AR-1242-4	5.833	5.003	31410	122074	1.004	8.672 #
20)	L4 AR-1242-5	6.592	5.508	368767	80915	11.954	4.842 #
21)	L5 AR-1248-1	5.655	4.725	111028	119937	3.268	8.653 #
22)	L5 AR-1248-2	5.915	4.955	24081	98122	0.460	4.940 #
23)	L5 AR-1248-3	6.137	5.003	74580	122074	1.309	5.854 #
24)	L5 AR-1248-4	6.538	5.160	158912	47010	2.914	1.912 #
25)	L5 AR-1248-5	6.592	5.535	368767	13743	6.492	0.633 #
26)	L6 AR-1254-1	6.515	5.508	143424	80915	2.179	2.284
27)	L6 AR-1254-2	6.723	5.673	165647	163381	1.721	5.108 #
28)	L6 AR-1254-3	7.101	6.078	201273	729000	2.252	14.912 #
29)	L6 AR-1254-4	7.384	6.297	14381	62491	0.262	2.385 #
30)	L6 AR-1254-5	7.828	6.706	698528	92434	10.374	2.193 #
31)	L7 AR-1260-1	7.254	6.198	51254	74923	0.715	2.124 #
32)	L7 AR-1260-2	7.538	6.383	397146	96431	5.034	2.406 #
33)	L7 AR-1260-3	7.869	6.536	15826	71906	0.290	1.902 #
34)	L7 AR-1260-4	8.104	7.009	41634	38295	0.664	1.325 #
35)	L7 AR-1260-5	8.426	7.252	289834	173290	2.708	2.880
36)	L8 AR-1262-1	7.869	6.743	15826	36060	0.175	0.769 #
37)	L8 AR-1262-2	8.426	7.009	289834	38295	2.174	0.922 #
38)	L8 AR-1262-3	8.723	7.535	91179	177461	0.927	5.555 #
39)	L8 AR-1262-4	8.837	7.605	477043	106977	5.955	1.946 #
40)	L8 AR-1262-5	9.535	8.094	71012	25728	1.616	1.024 #
41)	L9 AR-1268-1	8.723	7.535	91179	177461	0.529	1.990 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102723\
Data File : PO099041.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2023 17:46
Operator : YP/AJ
Sample : 05092-03
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 00:41:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.837	7.605	477043	106977	3.042	1.355 #
43)	L9 AR-1268-3	9.068	7.800	51608	42451	0.368	0.603 #
44)	L9 AR-1268-4	9.535	8.094	71012	25728	1.416	0.934 #
45)	L9 AR-1268-5	9.933	8.377	30269	17946	0.074	0.094 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102723\
Data File : PO099041.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2023 17:46
Operator : YP/AJ
Sample : 05092-03
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 00:41:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.475 min
Delta R.T.: -0.003 min
Response: 20495828
Conc: 10.37 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.630 min
Delta R.T.: -0.002 min
Response: 7677582
Conc: 10.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.277 min
Delta R.T.: -0.003 min
Response: 7884366
Conc: 6.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.627 min
Delta R.T.: -0.003 min
Response: 3635008
Conc: 6.78 ng/ml



#3 AR-1016-1


R.T.: 5.655 min
Delta R.T.: 0.004 min
Response: 111028
Conc: 2.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.725 min
Delta R.T.: 0.011 min
Response: 119937
Conc: 5.36 ng/ml




#4 AR-1016-2

R.T.: 5.655 min
Delta R.T.: -0.019 min
Response: 111028
Conc: 1.34 ng/ml



#4 AR-1016-2


R.T.: 4.734 min
Delta R.T.: 0.003 min
Response: 58477
Conc: 1.87 ng/ml



#5 AR-1016-3


R.T.: 5.736 min
Delta R.T.: 0.000 min
Response: 151393
Conc: 2.94 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.898 min
Delta R.T.: -0.009 min
Response: 111200
Conc: 6.57 ng/ml




#6 AR-1016-4

R.T.: 5.833 min
Delta R.T.: -0.002 min
Response: 31410
Conc: 0.77 ng/ml



#6 AR-1016-4


R.T.: 4.955 min
Delta R.T.: 0.005 min
Response: 98122
Conc: 6.79 ng/ml



#7 AR-1016-5


R.T.: 6.137 min
Delta R.T.: 0.006 min
Response: 74580
Conc: 1.76 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.160 min
Delta R.T.: -0.002 min
Response: 47010
Conc: 2.52 ng/ml




#8 AR-1221-1

R.T.: 4.684 min
Delta R.T.: -0.001 min
Response: 1097178
Conc: 44.19 ng/ml



#8 AR-1221-1


R.T.: 3.812 min
Delta R.T.: -0.033 min
Response: 24090
Conc: 2.59 ng/ml



#9 AR-1221-2


R.T.: 4.784 min
Delta R.T.: 0.011 min
Response: 660106
Conc: 35.68 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.910 min
Delta R.T.: -0.020 min
Response: 613635
Conc: 94.64 ng/ml




#10 AR-1221-3

R.T.: 4.849 min
Delta R.T.: 0.000 min
Response: 69918
Conc: 1.29 ng/ml



#10 AR-1221-3


R.T.: 4.005 min
Delta R.T.: 0.000 min
Response: 27631
Conc: 1.37 ng/ml



#11 AR-1232-1


R.T.: 4.849 min
Delta R.T.: 0.000 min
Response: 69918
Conc: 1.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.005 min
Delta R.T.: 0.000 min
Response: 27631
Conc: 1.61 ng/ml




#12 AR-1232-2

R.T.: 5.404 min
Delta R.T.: 0.023 min
Response: 83846
Conc: 3.55 ng/ml



#12 AR-1232-2


R.T.: 4.734 min
Delta R.T.: 0.004 min
Response: 58477
Conc: 3.82 ng/ml



#13 AR-1232-3


R.T.: 5.655 min
Delta R.T.: -0.017 min
Response: 111028
Conc: 2.81 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.898 min
Delta R.T.: -0.008 min
Response: 111200
Conc: 13.90 ng/ml




#14 AR-1232-4

R.T.: 5.833 min
Delta R.T.: 0.000 min
Response: 31410
Conc: 1.65 ng/ml



#14 AR-1232-4


R.T.: 5.003 min
Delta R.T.: 0.013 min
Response: 122074
Conc: 15.87 ng/ml



#15 AR-1232-5


R.T.: 5.915 min
Delta R.T.: -0.010 min
Response: 24081
Conc: 1.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.160 min
Delta R.T.: -0.001 min
Response: 47010
Conc: 5.53 ng/ml




#16 AR-1242-1

R.T.: 5.655 min
Delta R.T.: 0.004 min
Response: 111028
Conc: 2.56 ng/ml



#16 AR-1242-1


R.T.: 4.725 min
Delta R.T.: 0.013 min
Response: 119937
Conc: 6.83 ng/ml



#17 AR-1242-2


R.T.: 5.655 min
Delta R.T.: -0.018 min
Response: 111028
Conc: 1.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.734 min
Delta R.T.: 0.004 min
Response: 58477
Conc: 2.41 ng/ml




#18 AR-1242-3

R.T.: 5.736 min
Delta R.T.: 0.000 min
Response: 151393
Conc: 3.76 ng/ml



#18 AR-1242-3


R.T.: 4.898 min
Delta R.T.: -0.008 min
Response: 111200
Conc: 8.67 ng/ml



#19 AR-1242-4


R.T.: 5.833 min
Delta R.T.: -0.001 min
Response: 31410
Conc: 1.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.003 min
Delta R.T.: 0.014 min
Response: 122074
Conc: 8.67 ng/ml




#20 AR-1242-5

R.T.: 6.592 min
Delta R.T.: 0.018 min
Response: 368767
Conc: 11.95 ng/ml



#20 AR-1242-5


R.T.: 5.508 min
Delta R.T.: -0.004 min
Response: 80915
Conc: 4.84 ng/ml



#21 AR-1248-1


R.T.: 5.655 min
Delta R.T.: 0.005 min
Response: 111028
Conc: 3.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.725 min
Delta R.T.: 0.013 min
Response: 119937
Conc: 8.65 ng/ml




#22 AR-1248-2

R.T.: 5.915 min
Delta R.T.: -0.010 min
Response: 24081
Conc: 0.46 ng/ml



#22 AR-1248-2


R.T.: 4.955 min
Delta R.T.: 0.006 min
Response: 98122
Conc: 4.94 ng/ml



#23 AR-1248-3


R.T.: 6.137 min
Delta R.T.: 0.006 min
Response: 74580
Conc: 1.31 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.003 min
Delta R.T.: 0.013 min
Response: 122074
Conc: 5.85 ng/ml




#24 AR-1248-4

R.T.: 6.538 min
Delta R.T.: 0.003 min
Response: 158912
Conc: 2.91 ng/ml



#24 AR-1248-4


R.T.: 5.160 min
Delta R.T.: -0.001 min
Response: 47010
Conc: 1.91 ng/ml



#25 AR-1248-5


R.T.: 6.592 min
Delta R.T.: 0.018 min
Response: 368767
Conc: 6.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.535 min
Delta R.T.: -0.019 min
Response: 13743
Conc: 0.63 ng/ml




#26 AR-1254-1

R.T.: 6.515 min
Delta R.T.: 0.006 min
Response: 143424
Conc: 2.18 ng/ml



#26 AR-1254-1


R.T.: 5.508 min
Delta R.T.: -0.004 min
Response: 80915
Conc: 2.28 ng/ml



#27 AR-1254-2


R.T.: 6.723 min
Delta R.T.: -0.005 min
Response: 165647
Conc: 1.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.673 min
Delta R.T.: 0.013 min
Response: 163381
Conc: 5.11 ng/ml




#28 AR-1254-3

R.T.: 7.101 min
Delta R.T.: 0.007 min
Response: 201273
Conc: 2.25 ng/ml



#28 AR-1254-3


R.T.: 6.078 min
Delta R.T.: 0.015 min
Response: 729000
Conc: 14.91 ng/ml



#29 AR-1254-4


R.T.: 7.384 min
Delta R.T.: 0.004 min
Response: 14381
Conc: 0.26 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.297 min
Delta R.T.: 0.006 min
Response: 62491
Conc: 2.38 ng/ml




#30 AR-1254-5

R.T.: 7.828 min
Delta R.T.: 0.028 min
Response: 698528
Conc: 10.37 ng/ml



#30 AR-1254-5


R.T.: 6.706 min
Delta R.T.: -0.003 min
Response: 92434
Conc: 2.19 ng/ml



#31 AR-1260-1


R.T.: 7.254 min
Delta R.T.: -0.006 min
Response: 51254
Conc: 0.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.198 min
Delta R.T.: 0.003 min
Response: 74923
Conc: 2.12 ng/ml




#32 AR-1260-2

R.T.: 7.538 min
Delta R.T.: 0.021 min
Response: 397146
Conc: 5.03 ng/ml



#32 AR-1260-2


R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 96431
Conc: 2.41 ng/ml



#33 AR-1260-3


R.T.: 7.869 min
Delta R.T.: -0.008 min
Response: 15826
Conc: 0.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.536 min
Delta R.T.: 0.000 min
Response: 71906
Conc: 1.90 ng/ml




#34 AR-1260-4

R.T.: 8.104 min
Delta R.T.: 0.000 min
Response: 41634
Conc: 0.66 ng/ml



#34 AR-1260-4


R.T.: 7.009 min
Delta R.T.: 0.000 min
Response: 38295
Conc: 1.33 ng/ml



#35 AR-1260-5


R.T.: 8.426 min
Delta R.T.: -0.002 min
Response: 289834
Conc: 2.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.252 min
Delta R.T.: 0.002 min
Response: 173290
Conc: 2.88 ng/ml




#36 AR-1262-1

R.T.: 7.869 min
Delta R.T.: -0.007 min
Response: 15826
Conc: 0.17 ng/ml



#36 AR-1262-1


R.T.: 6.743 min
Delta R.T.: -0.004 min
Response: 36060
Conc: 0.77 ng/ml



#37 AR-1262-2


R.T.: 8.426 min
Delta R.T.: -0.001 min
Response: 289834
Conc: 2.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.009 min
Delta R.T.: 0.002 min
Response: 38295
Conc: 0.92 ng/ml




#38 AR-1262-3

R.T.: 8.723 min
Delta R.T.: -0.022 min
Response: 91179
Conc: 0.93 ng/ml



#38 AR-1262-3


R.T.: 7.535 min
Delta R.T.: 0.003 min
Response: 177461
Conc: 5.55 ng/ml



#39 AR-1262-4


R.T.: 8.837 min
Delta R.T.: 0.001 min
Response: 477043
Conc: 5.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.605 min
Delta R.T.: 0.010 min
Response: 106977
Conc: 1.95 ng/ml




#40 AR-1262-5

R.T.: 9.535 min
Delta R.T.: 0.030 min
Response: 71012
Conc: 1.62 ng/ml



#40 AR-1262-5


R.T.: 8.094 min
Delta R.T.: 0.004 min
Response: 25728
Conc: 1.02 ng/ml



#41 AR-1268-1


R.T.: 8.723 min
Delta R.T.: -0.021 min
Response: 91179
Conc: 0.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.535 min
Delta R.T.: 0.003 min
Response: 177461
Conc: 1.99 ng/ml




#42 AR-1268-2

R.T.: 8.837 min
Delta R.T.: -0.004 min
Response: 477043
Conc: 3.04 ng/ml



#42 AR-1268-2


R.T.: 7.605 min
Delta R.T.: 0.008 min
Response: 106977
Conc: 1.35 ng/ml



#43 AR-1268-3


R.T.: 9.068 min
Delta R.T.: -0.004 min
Response: 51608
Conc: 0.37 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.800 min
Delta R.T.: -0.002 min
Response: 42451
Conc: 0.60 ng/ml




#44 AR-1268-4

R.T.: 9.535 min
Delta R.T.: 0.028 min
Response: 71012
Conc: 1.42 ng/ml



#44 AR-1268-4


R.T.: 8.094 min
Delta R.T.: 0.004 min
Response: 25728
Conc: 0.93 ng/ml



#45 AR-1268-5

R.T.: 9.933 min
Delta R.T.: 0.000 min
Response: 30269
Conc: 0.07 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.377 min
Delta R.T.: -0.002 min
Response: 17946
Conc: 0.09 ng/ml