

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0102723\
 Data File : P0099084.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2023 07:49
 Operator : YP/AJ
 Sample : 05093-01 10X
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 23:12:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.473	3.627	4459367	1489779	2.257	2.070
2)	SA Decachlor...	10.275	8.625	3328103	1241718	2.938	2.314
Target Compounds							
3)	L1 AR-1016-1	5.650	4.671f	99489	319612	1.807	14.272 #
4)	L1 AR-1016-2	5.674	4.762	215231	147700	2.603	4.725 #
5)	L1 AR-1016-3	5.735	0.000	18457	0	0.359	N.D. #
6)	L1 AR-1016-4	5.821	0.000	21665	0	0.533	N.D. #
7)	L1 AR-1016-5	6.147	5.162	27796	83431	0.655	4.466 #
8)	L2 AR-1221-1	4.657	3.856	1833717	73055	73.856	7.864 #
9)	L2 AR-1221-2	4.787	3.931	2129214	95032	115.097	14.657 #
10)	L2 AR-1221-3	4.858	3.972	645800	94312	11.920	4.663 #
11)	L3 AR-1232-1	4.858	3.972	645800	94312	14.117	5.502 #
12)	L3 AR-1232-2	5.386	4.762	44702	147700	1.893	9.641 #
13)	L3 AR-1232-3	5.674	0.000	215231	0	5.450	N.D. #
14)	L3 AR-1232-4	5.821	0.000	21665	0	1.136	N.D. #
15)	L3 AR-1232-5	5.919	5.162	74988	83431	4.502	9.809 #
16)	L4 AR-1242-1	5.650	4.671f	99489	319612	2.298	18.191 #
17)	L4 AR-1242-2	5.674	4.762	215231	147700	3.361	6.095 #
18)	L4 AR-1242-3	5.735	0.000	18457	0	0.459	N.D. #
19)	L4 AR-1242-4	5.821	0.000	21665	0	0.692	N.D. #
20)	L4 AR-1242-5	6.571	5.525	46616	64491	1.511	3.859 #
21)	L5 AR-1248-1	5.650	4.671f	99489	319612	2.928	23.060 #
22)	L5 AR-1248-2	5.919	0.000	74988	0	1.433	N.D. #
23)	L5 AR-1248-3	6.147	0.000	27796	0	0.488	N.D. #
24)	L5 AR-1248-4	6.532	5.162	72193	83431	1.324	3.394 #
25)	L5 AR-1248-5	6.571	5.525	46616	64491	0.821	2.969 #
26)	L6 AR-1254-1	6.505	5.525	164271	64491	2.495	1.820 #
27)	L6 AR-1254-2	6.724	5.656	292735	95497	3.041	2.986
28)	L6 AR-1254-3	7.106	6.062	486626	185596	5.444	3.796 #
29)	L6 AR-1254-4	7.384	6.289	84361	79086	1.540	3.018 #
30)	L6 AR-1254-5	7.805	6.707	290853	290967	4.320	6.902 #
31)	L7 AR-1260-1	7.258	6.197	59628	181138	0.832	5.135 #
32)	L7 AR-1260-2	7.524	6.372	79989	50487	1.014	1.259
33)	L7 AR-1260-3	7.847	0.000	103754	0	1.900	N.D. #
34)	L7 AR-1260-4	8.104	6.996	125090	152788	1.996	5.287 #
35)	L7 AR-1260-5	8.422	7.255	157591	129712	1.473	2.155 #
36)	L8 AR-1262-1	7.847	6.707f	103754	290967	1.146	6.201 #
37)	L8 AR-1262-2	8.422	6.996	157591	152788	1.182	3.680 #
38)	L8 AR-1262-3	8.733	7.499	23152	155079	0.235	4.854 #
39)	L8 AR-1262-4	8.837	7.594	124157	99596	1.550	1.812
40)	L8 AR-1262-5	9.504	8.055f	86902	176059	1.978	7.010 #
41)	L9 AR-1268-1	8.733	7.499	23152	155079	0.134	1.739 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102723\
Data File : PO099084.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2023 07:49
Operator : YP/AJ
Sample : 05093-01 10X
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 23:12:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.837	7.594	124157	99596	0.792	1.261 #
43)	L9 AR-1268-3	9.069	7.793	83589	74275	0.596	1.055 #
44)	L9 AR-1268-4	9.504	8.055f	86902	176059	1.733	6.388 #
45)	L9 AR-1268-5	9.930	8.374	-77385	322511	N.D.	1.692 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102723\
Data File : PO099084.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2023 07:49
Operator : YP/AJ
Sample : 05093-01 10X
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 23:12:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.473 min
Delta R.T.: -0.005 min
Response: 4459367
Conc: 2.26 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.627 min
Delta R.T.: -0.005 min
Response: 1489779
Conc: 2.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.275 min
Delta R.T.: -0.005 min
Response: 3328103
Conc: 2.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.625 min
Delta R.T.: -0.005 min
Response: 1241718
Conc: 2.31 ng/ml



#3 AR-1016-1


R.T.: 5.650 min
Delta R.T.: 0.000 min
Response: 99489
Conc: 1.81 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.671 min
Delta R.T.: -0.042 min
Response: 319612
Conc: 14.27 ng/ml




#4 AR-1016-2

R.T.: 5.674 min
Delta R.T.: 0.000 min
Response: 215231
Conc: 2.60 ng/ml



#4 AR-1016-2


R.T.: 4.762 min
Delta R.T.: 0.030 min
Response: 147700
Conc: 4.73 ng/ml



#5 AR-1016-3


R.T.: 5.735 min
Delta R.T.: 0.000 min
Response: 18457
Conc: 0.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.908 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.821 min
Delta R.T.: -0.013 min
Response: 21665
Conc: 0.53 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.950 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.147 min
Delta R.T.: 0.016 min
Response: 27796
Conc: 0.65 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.162 min
Delta R.T.: 0.000 min
Response: 83431
Conc: 4.47 ng/ml




#8 AR-1221-1

R.T.: 4.657 min
Delta R.T.: -0.028 min
Response: 1833717
Conc: 73.86 ng/ml



#8 AR-1221-1


R.T.: 3.856 min
Delta R.T.: 0.011 min
Response: 73055
Conc: 7.86 ng/ml



#9 AR-1221-2


R.T.: 4.787 min
Delta R.T.: 0.014 min
Response: 2129214
Conc: 115.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.931 min
Delta R.T.: 0.001 min
Response: 95032
Conc: 14.66 ng/ml




#10 AR-1221-3

R.T.: 4.858 min
Delta R.T.: 0.008 min
Response: 645800
Conc: 11.92 ng/ml



#10 AR-1221-3


R.T.: 3.972 min
Delta R.T.: -0.034 min
Response: 94312
Conc: 4.66 ng/ml



#11 AR-1232-1


R.T.: 4.858 min
Delta R.T.: 0.009 min
Response: 645800
Conc: 14.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.972 min
Delta R.T.: -0.033 min
Response: 94312
Conc: 5.50 ng/ml




#12 AR-1232-2

R.T.: 5.386 min
Delta R.T.: 0.005 min
Response: 44702
Conc: 1.89 ng/ml



#12 AR-1232-2


R.T.: 4.762 min
Delta R.T.: 0.031 min
Response: 147700
Conc: 9.64 ng/ml



#13 AR-1232-3


R.T.: 5.674 min
Delta R.T.: 0.002 min
Response: 215231
Conc: 5.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.906 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.821 min
Delta R.T.: -0.011 min
Response: 21665
Conc: 1.14 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 4.990 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.919 min
Delta R.T.: -0.005 min
Response: 74988
Conc: 4.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.162 min
Delta R.T.: 0.000 min
Response: 83431
Conc: 9.81 ng/ml




#16 AR-1242-1

R.T.: 5.650 min
Delta R.T.: 0.000 min
Response: 99489
Conc: 2.30 ng/ml



#16 AR-1242-1


R.T.: 4.671 min
Delta R.T.: -0.041 min
Response: 319612
Conc: 18.19 ng/ml



#17 AR-1242-2


R.T.: 5.674 min
Delta R.T.: 0.001 min
Response: 215231
Conc: 3.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.762 min
Delta R.T.: 0.032 min
Response: 147700
Conc: 6.09 ng/ml




#18 AR-1242-3

R.T.: 5.735 min
Delta R.T.: 0.000 min
Response: 18457
Conc: 0.46 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.906 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.821 min
Delta R.T.: -0.013 min
Response: 21665
Conc: 0.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 4.989 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.571 min
Delta R.T.: -0.002 min
Response: 46616
Conc: 1.51 ng/ml



#20 AR-1242-5


R.T.: 5.525 min
Delta R.T.: 0.012 min
Response: 64491
Conc: 3.86 ng/ml



#21 AR-1248-1


R.T.: 5.650 min
Delta R.T.: 0.000 min
Response: 99489
Conc: 2.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.671 min
Delta R.T.: -0.041 min
Response: 319612
Conc: 23.06 ng/ml




#22 AR-1248-2

R.T.: 5.919 min
Delta R.T.: -0.005 min
Response: 74988
Conc: 1.43 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 4.949 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.147 min
Delta R.T.: 0.017 min
Response: 27796
Conc: 0.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.990 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.532 min
Delta R.T.: -0.003 min
Response: 72193
Conc: 1.32 ng/ml



#24 AR-1248-4


R.T.: 5.162 min
Delta R.T.: 0.000 min
Response: 83431
Conc: 3.39 ng/ml



#25 AR-1248-5


R.T.: 6.571 min
Delta R.T.: -0.002 min
Response: 46616
Conc: 0.82 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.525 min
Delta R.T.: -0.029 min
Response: 64491
Conc: 2.97 ng/ml




#26 AR-1254-1

R.T.: 6.505 min
Delta R.T.: -0.004 min
Response: 164271
Conc: 2.50 ng/ml



#26 AR-1254-1


R.T.: 5.525 min
Delta R.T.: 0.013 min
Response: 64491
Conc: 1.82 ng/ml



#27 AR-1254-2


R.T.: 6.724 min
Delta R.T.: -0.004 min
Response: 292735
Conc: 3.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.656 min
Delta R.T.: -0.004 min
Response: 95497
Conc: 2.99 ng/ml




#28 AR-1254-3

R.T.: 7.106 min
Delta R.T.: 0.012 min
Response: 486626
Conc: 5.44 ng/ml



#28 AR-1254-3


R.T.: 6.062 min
Delta R.T.: -0.001 min
Response: 185596
Conc: 3.80 ng/ml



#29 AR-1254-4


R.T.: 7.384 min
Delta R.T.: 0.004 min
Response: 84361
Conc: 1.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.289 min
Delta R.T.: -0.003 min
Response: 79086
Conc: 3.02 ng/ml




#30 AR-1254-5

R.T.: 7.805 min
Delta R.T.: 0.005 min
Response: 290853
Conc: 4.32 ng/ml



#30 AR-1254-5


R.T.: 6.707 min
Delta R.T.: -0.002 min
Response: 290967
Conc: 6.90 ng/ml



#31 AR-1260-1


R.T.: 7.258 min
Delta R.T.: -0.002 min
Response: 59628
Conc: 0.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.197 min
Delta R.T.: 0.002 min
Response: 181138
Conc: 5.13 ng/ml




#32 AR-1260-2

R.T.: 7.524 min
Delta R.T.: 0.007 min
Response: 79989
Conc: 1.01 ng/ml



#32 AR-1260-2


R.T.: 6.372 min
Delta R.T.: -0.011 min
Response: 50487
Conc: 1.26 ng/ml



#33 AR-1260-3


R.T.: 7.847 min
Delta R.T.: -0.030 min
Response: 103754
Conc: 1.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.537 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.104 min
Delta R.T.: 0.000 min
Response: 125090
Conc: 2.00 ng/ml



#34 AR-1260-4


R.T.: 6.996 min
Delta R.T.: -0.012 min
Response: 152788
Conc: 5.29 ng/ml



#35 AR-1260-5


R.T.: 8.422 min
Delta R.T.: -0.006 min
Response: 157591
Conc: 1.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.255 min
Delta R.T.: 0.005 min
Response: 129712
Conc: 2.16 ng/ml




#36 AR-1262-1

R.T.: 7.847 min
Delta R.T.: -0.029 min
Response: 103754
Conc: 1.15 ng/ml



#36 AR-1262-1


R.T.: 6.707 min
Delta R.T.: -0.040 min
Response: 290967
Conc: 6.20 ng/ml



#37 AR-1262-2


R.T.: 8.422 min
Delta R.T.: -0.005 min
Response: 157591
Conc: 1.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.996 min
Delta R.T.: -0.010 min
Response: 152788
Conc: 3.68 ng/ml




#38 AR-1262-3

R.T.: 8.733 min
Delta R.T.: -0.012 min
Response: 23152
Conc: 0.24 ng/ml



#38 AR-1262-3


R.T.: 7.499 min
Delta R.T.: -0.032 min
Response: 155079
Conc: 4.85 ng/ml



#39 AR-1262-4


R.T.: 8.837 min
Delta R.T.: 0.000 min
Response: 124157
Conc: 1.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.594 min
Delta R.T.: -0.002 min
Response: 99596
Conc: 1.81 ng/ml




#40 AR-1262-5

R.T.: 9.504 min
Delta R.T.: -0.002 min
Response: 86902
Conc: 1.98 ng/ml



#40 AR-1262-5


R.T.: 8.055 min
Delta R.T.: -0.036 min
Response: 176059
Conc: 7.01 ng/ml



#41 AR-1268-1


R.T.: 8.733 min
Delta R.T.: -0.011 min
Response: 23152
Conc: 0.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.499 min
Delta R.T.: -0.033 min
Response: 155079
Conc: 1.74 ng/ml




#42 AR-1268-2

R.T.: 8.837 min
Delta R.T.: -0.004 min
Response: 124157
Conc: 0.79 ng/ml



#42 AR-1268-2


R.T.: 7.594 min
Delta R.T.: -0.003 min
Response: 99596
Conc: 1.26 ng/ml



#43 AR-1268-3


R.T.: 9.069 min
Delta R.T.: -0.003 min
Response: 83589
Conc: 0.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.793 min
Delta R.T.: -0.009 min
Response: 74275
Conc: 1.05 ng/ml




#44 AR-1268-4

R.T.: 9.504 min
Delta R.T.: -0.003 min
Response: 86902
Conc: 1.73 ng/ml



#44 AR-1268-4


R.T.: 8.055 min
Delta R.T.: -0.036 min
Response: 176059
Conc: 6.39 ng/ml



#45 AR-1268-5

R.T.: 9.930 min
Delta R.T.: -0.003 min
Response: -77385
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.374 min
Delta R.T.: -0.004 min
Response: 322511
Conc: 1.69 ng/ml