

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102723\
 Data File : PO099029.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2023 11:25
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 15:24:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.474	3.628	92207355	36119953	46.665	50.189
2) SA Decachlor...	10.277	8.628	57632390	26299905	50.873	49.018
Target Compounds						
21) L5 AR-1248-1	5.649	4.710	16728462	7145226	492.409	515.529
22) L5 AR-1248-2	5.923	4.946	24603783	10297123	470.136	518.392
23) L5 AR-1248-3	6.128	4.987	26679103	10785998	468.237	517.226
24) L5 AR-1248-4	6.533	5.159	26901565	12538091	493.320	510.021
25) L5 AR-1248-5	6.572	5.551	27154047	11462929	478.013	527.798



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102723\
Data File : PO099029.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2023 11:25
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 15:24:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.474 min
Delta R.T.: -0.004 min
Response: 92207355
Conc: 46.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.628 min
Delta R.T.: -0.003 min
Response: 36119953
Conc: 50.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.277 min
Delta R.T.: -0.003 min
Response: 57632390
Conc: 50.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.628 min
Delta R.T.: -0.003 min
Response: 26299905
Conc: 49.02 ng/ml



#21 AR-1248-1


R.T.: 5.649 min
Delta R.T.: -0.002 min
Response: 16728462
Conc: 492.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.710 min
Delta R.T.: -0.002 min
Response: 7145226
Conc: 515.53 ng/ml




#22 AR-1248-2

R.T.: 5.923 min
Delta R.T.: -0.001 min
Response: 24603783
Conc: 470.14 ng/ml



#22 AR-1248-2


R.T.: 4.946 min
Delta R.T.: -0.003 min
Response: 10297123
Conc: 518.39 ng/ml



#23 AR-1248-3


R.T.: 6.128 min
Delta R.T.: -0.002 min
Response: 26679103
Conc: 468.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.987 min
Delta R.T.: -0.003 min
Response: 10785998
Conc: 517.23 ng/ml




#24 AR-1248-4

R.T.: 6.533 min
Delta R.T.: -0.002 min
Response: 26901565
Conc: 493.32 ng/ml



#24 AR-1248-4


R.T.: 5.159 min
Delta R.T.: -0.003 min
Response: 12538091
Conc: 510.02 ng/ml



#25 AR-1248-5

R.T.: 6.572 min
Delta R.T.: -0.002 min
Response: 27154047
Conc: 478.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.551 min
Delta R.T.: -0.003 min
Response: 11462929
Conc: 527.80 ng/ml