

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103019\
 Data File : PO063310.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Oct 2019 18:40
 Operator : HP/AJ
 Sample : K5599-13
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 Z3-23A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 01:56:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 30 02:25:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.316	3.480	1235133	1182488	21.193	21.652
2)	SA Decachlor...	9.931	8.321	1064574	740673	18.296	17.325
Target Compounds							
27)	L6 AR-1254-2	0.000	5.490	0	99629	N.D.	36.941 #
28)	L6 AR-1254-3	0.000	5.862	0	25185	N.D.	5.866 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103019\
Data File : PO063310.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Oct 2019 18:40
Operator : HP/AJ
Sample : K5599-13
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
Z3-23A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 01:56:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102919.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 30 02:25:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.316 min
Delta R.T.: 0.000 min
Response: 1235133
Conc: 21.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
Z3-23A




#1 Tetrachloro-m-xylene

R.T.: 3.480 min
Delta R.T.: 0.002 min
Response: 1182488
Conc: 21.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.931 min
Delta R.T.: -0.003 min
Response: 1064574
Conc: 18.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.321 min
Delta R.T.: 0.000 min
Response: 740673
Conc: 17.32 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 6.539 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
Z3-23A




#27 AR-1254-2

R.T.: 5.490 min
Delta R.T.: 0.050 min
Response: 99629
Conc: 36.94 ng/ml



#28 AR-1254-3

R.T.: 0.000 min
Exp R.T.: 6.904 min
Response: 0
Conc: N.D.



#28 AR-1254-3

R.T.: 5.862 min
Delta R.T.: 0.030 min
Response: 25185
Conc: 5.87 ng/ml