

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103020\
 Data File : PO072908.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Oct 2020 11:54
 Operator : DD\AJ
 Sample : AR1221PO1030ICV500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 12:09:14 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 30 11:30:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.948	4.024	2573413	1804563	46.942	46.787
2)	SA Decachlor...	10.983	9.308	2104780	1811957	46.760	46.506
Target Compounds							
8)	L2 AR-1221-1	5.199	4.281	315497	219040	464.340	448.536
9)	L2 AR-1221-2	5.298	4.379	239853	169287	477.468	464.657
10)	L2 AR-1221-3	5.385	4.468	701293	510956	453.222	445.243



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103020\
Data File : PO072908.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Oct 2020 11:54
Operator : DD\AJ
Sample : AR1221P01030ICV500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 12:09:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102620.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 30 11:30:59 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.948 min
Delta R.T.: 0.000 min
Response: 2573413
Conc: 46.94 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.024 min
Delta R.T.: 0.000 min
Response: 1804563
Conc: 46.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.983 min
Delta R.T.: -0.003 min
Response: 2104780
Conc: 46.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.308 min
Delta R.T.: 0.000 min
Response: 1811957
Conc: 46.51 ng/ml



#8 AR-1221-1


R.T.: 5.199 min
Delta R.T.: -0.001 min
Response: 315497
Conc: 464.34 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 4.281 min
Delta R.T.: 0.000 min
Response: 219040
Conc: 448.54 ng/ml




#9 AR-1221-2

R.T.: 5.298 min
Delta R.T.: -0.001 min
Response: 239853
Conc: 477.47 ng/ml



#9 AR-1221-2


R.T.: 4.379 min
Delta R.T.: 0.000 min
Response: 169287
Conc: 464.66 ng/ml



#10 AR-1221-3

R.T.: 5.385 min
Delta R.T.: -0.001 min
Response: 701293
Conc: 453.22 ng/ml

Instrument :
ECD_O
ClientSampleId :



#10 AR-1221-3

R.T.: 4.468 min
Delta R.T.: 0.000 min
Response: 510956
Conc: 445.24 ng/ml