

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103020\
 Data File : PO072911.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Oct 2020 13:26
 Operator : DD\AJ
 Sample : AR1221CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 04:55:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 30 12:29:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.950	4.025	2595356	1796017	56.290	51.245
2) SA Decachlor...	10.985	9.310	2101173	1839686	56.873	55.165
Target Compounds						
8) L2 AR-1221-1	5.200	4.283	315450	219310	464.271	449.090
9) L2 AR-1221-2	5.301	4.381	237296	170444	472.379	467.833
10) L2 AR-1221-3	5.387	4.469	699638	508572	452.153	443.165



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103020\
Data File : PO072911.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Oct 2020 13:26
Operator : DD\AJ
Sample : AR1221CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 04:55:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102620.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 30 12:29:20 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.950 min
Delta R.T.: 0.001 min
Response: 2595356
Conc: 56.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.025 min
Delta R.T.: 0.005 min
Response: 1796017
Conc: 51.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.985 min
Delta R.T.: 0.000 min
Response: 2101173
Conc: 56.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.310 min
Delta R.T.: -0.003 min
Response: 1839686
Conc: 55.16 ng/ml



#8 AR-1221-1


R.T.: 5.200 min
Delta R.T.: 0.002 min
Response: 315450
Conc: 464.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221CCC500




#8 AR-1221-1

R.T.: 4.283 min
Delta R.T.: 0.000 min
Response: 219310
Conc: 449.09 ng/ml




#9 AR-1221-2

R.T.: 5.301 min
Delta R.T.: 0.001 min
Response: 237296
Conc: 472.38 ng/ml



#9 AR-1221-2


R.T.: 4.381 min
Delta R.T.: 0.000 min
Response: 170444
Conc: 467.83 ng/ml



#10 AR-1221-3

R.T.: 5.387 min
Delta R.T.: 0.000 min
Response: 699638
Conc: 452.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221CCC500



#10 AR-1221-3

R.T.: 4.469 min
Delta R.T.: 0.000 min
Response: 508572
Conc: 443.17 ng/ml