

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103023\
 Data File : PO099119.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2023 08:39
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 11:22:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.476	3.618	73665	105859	0.037	0.147 #
2)	SA Decachlor...	10.278	8.614	20091	5712	0.018	0.011 #
Target Compounds							
3)	L1 AR-1016-1	5.647	4.720	22063	89342	0.401	3.990 #
4)	L1 AR-1016-2	5.658	4.742	35586	11163	0.430	0.357
5)	L1 AR-1016-3	5.735	4.907	51984	29789	1.010	1.759 #
6)	L1 AR-1016-4	5.828	4.935	97604	23578	2.401	1.631 #
7)	L1 AR-1016-5	6.131	5.143	211521	87037	4.984	4.659
8)	L2 AR-1221-1	4.687	3.853	87392	17087	3.520	1.839 #
9)	L2 AR-1221-2	4.759	3.933	400140	12605	21.630	1.944 #
10)	L2 AR-1221-3	4.857	4.009	107196	27393	1.979	1.354 #
11)	L3 AR-1232-1	4.857	4.009	107196	27393	2.343	1.598 #
12)	L3 AR-1232-2	5.377	4.720	33067	89342	1.400	5.832 #
13)	L3 AR-1232-3	5.658	4.907	35586	29789	0.901	3.723 #
14)	L3 AR-1232-4	5.828	4.981	97604	122345	5.116	15.902 #
15)	L3 AR-1232-5	5.912	5.143	105971	87037	6.361	10.233 #
16)	L4 AR-1242-1	5.647	4.720	22063	89342	0.510	5.085 #
17)	L4 AR-1242-2	5.658	4.720	35586	89342	0.556	3.687 #
18)	L4 AR-1242-3	5.735	4.907	51984	29789	1.292	2.324 #
19)	L4 AR-1242-4	5.828	4.981	97604	122345	3.119	8.691 #
20)	L4 AR-1242-5	6.581	5.543	740590	131750	24.007	7.884 #
21)	L5 AR-1248-1	5.647	4.720	22063	89342	0.649	6.446 #
22)	L5 AR-1248-2	5.912	4.935	105971	23578	2.025	1.187 #
23)	L5 AR-1248-3	6.131	4.981	211521	122345	3.712	5.867 #
24)	L5 AR-1248-4	6.532	5.143	647379	87037	11.872	3.540 #
25)	L5 AR-1248-5	6.581	5.561	740590	29930	13.037	1.378 #
26)	L6 AR-1254-1	6.532	5.543	647379	131750	9.835	3.718 #
27)	L6 AR-1254-2	6.744	5.663	671665	168074	6.977	5.255
28)	L6 AR-1254-3	7.101	6.059	264500	302552	2.959	6.189 #
29)	L6 AR-1254-4	7.392	6.270	55735	50404	1.017	1.923 #
30)	L6 AR-1254-5	7.798	6.669f	263007	33781	3.906	0.801 #
31)	L7 AR-1260-1	7.280	6.172	66782	245955	0.932	6.972 #
32)	L7 AR-1260-2	7.516	6.371	376696	115906	4.775	2.891 #
33)	L7 AR-1260-3	7.868	6.545	152950	33193	2.801	0.878 #
34)	L7 AR-1260-4	8.106	7.018	382698	186070	6.108	6.439
35)	L7 AR-1260-5	8.430	7.250	339452	28212	3.172	0.469 #
36)	L8 AR-1262-1	7.868	6.780	152950	396847	1.690	8.458 #
37)	L8 AR-1262-2	8.430	7.018	339452	186070	2.546	4.482 #
38)	L8 AR-1262-3	8.771	7.522	-217954	90729	N.D.	2.840 #
39)	L8 AR-1262-4	8.825	7.617	210167	111591	2.623	2.030
40)	L8 AR-1262-5	9.509	8.049f	36433	27686	0.829	1.102 #
42)	L9 AR-1268-2	8.856	7.617	264037	111591	1.684	1.413

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103023\
 Data File : PO099119.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2023 08:39
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 11:22:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	9.074	7.840f	133800	125427	0.954	1.782 #
44) L9	AR-1268-4	9.509	8.049f	36433	27686	0.727	1.005 #
45) L9	AR-1268-5	9.929	8.372	364755	183894	0.891	0.965



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103023\
Data File : PO099119.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2023 08:39
Operator : YP/AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 11:22:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.476 min
Delta R.T.: -0.002 min
Response: 73665
Conc: 0.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.618 min
Delta R.T.: -0.014 min
Response: 105859
Conc: 0.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.278 min
Delta R.T.: -0.002 min
Response: 20091
Conc: 0.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.614 min
Delta R.T.: -0.016 min
Response: 5712
Conc: 0.01 ng/ml



#3 AR-1016-1


R.T.: 5.647 min
Delta R.T.: -0.004 min
Response: 22063
Conc: 0.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.720 min
Delta R.T.: 0.006 min
Response: 89342
Conc: 3.99 ng/ml




#4 AR-1016-2

R.T.: 5.658 min
Delta R.T.: -0.016 min
Response: 35586
Conc: 0.43 ng/ml



#4 AR-1016-2


R.T.: 4.742 min
Delta R.T.: 0.011 min
Response: 11163
Conc: 0.36 ng/ml



#5 AR-1016-3


R.T.: 5.735 min
Delta R.T.: -0.001 min
Response: 51984
Conc: 1.01 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.907 min
Delta R.T.: 0.000 min
Response: 29789
Conc: 1.76 ng/ml




#6 AR-1016-4

R.T.: 5.828 min
Delta R.T.: -0.007 min
Response: 97604
Conc: 2.40 ng/ml



#6 AR-1016-4


R.T.: 4.935 min
Delta R.T.: -0.015 min
Response: 23578
Conc: 1.63 ng/ml



#7 AR-1016-5


R.T.: 6.131 min
Delta R.T.: 0.000 min
Response: 211521
Conc: 4.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.143 min
Delta R.T.: -0.019 min
Response: 87037
Conc: 4.66 ng/ml




#8 AR-1221-1

R.T.: 4.687 min
Delta R.T.: 0.002 min
Response: 87392
Conc: 3.52 ng/ml



#8 AR-1221-1


R.T.: 3.853 min
Delta R.T.: 0.009 min
Response: 17087
Conc: 1.84 ng/ml



#9 AR-1221-2


R.T.: 4.759 min
Delta R.T.: -0.013 min
Response: 400140
Conc: 21.63 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.933 min
Delta R.T.: 0.003 min
Response: 12605
Conc: 1.94 ng/ml




#10 AR-1221-3

R.T.: 4.857 min
Delta R.T.: 0.007 min
Response: 107196
Conc: 1.98 ng/ml



#10 AR-1221-3


R.T.: 4.009 min
Delta R.T.: 0.003 min
Response: 27393
Conc: 1.35 ng/ml



#11 AR-1232-1


R.T.: 4.857 min
Delta R.T.: 0.008 min
Response: 107196
Conc: 2.34 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.009 min
Delta R.T.: 0.004 min
Response: 27393
Conc: 1.60 ng/ml




#12 AR-1232-2

R.T.: 5.377 min
Delta R.T.: -0.004 min
Response: 33067
Conc: 1.40 ng/ml



#12 AR-1232-2


R.T.: 4.720 min
Delta R.T.: -0.011 min
Response: 89342
Conc: 5.83 ng/ml



#13 AR-1232-3


R.T.: 5.658 min
Delta R.T.: -0.014 min
Response: 35586
Conc: 0.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.907 min
Delta R.T.: 0.001 min
Response: 29789
Conc: 3.72 ng/ml




#14 AR-1232-4

R.T.: 5.828 min
Delta R.T.: -0.005 min
Response: 97604
Conc: 5.12 ng/ml



#14 AR-1232-4


R.T.: 4.981 min
Delta R.T.: -0.008 min
Response: 122345
Conc: 15.90 ng/ml



#15 AR-1232-5


R.T.: 5.912 min
Delta R.T.: -0.012 min
Response: 105971
Conc: 6.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.143 min
Delta R.T.: -0.019 min
Response: 87037
Conc: 10.23 ng/ml




#16 AR-1242-1

R.T.: 5.647 min
Delta R.T.: -0.004 min
Response: 22063
Conc: 0.51 ng/ml



#16 AR-1242-1


R.T.: 4.720 min
Delta R.T.: 0.008 min
Response: 89342
Conc: 5.08 ng/ml



#17 AR-1242-2


R.T.: 5.658 min
Delta R.T.: -0.015 min
Response: 35586
Conc: 0.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.720 min
Delta R.T.: -0.011 min
Response: 89342
Conc: 3.69 ng/ml




#18 AR-1242-3

R.T.: 5.735 min
Delta R.T.: 0.000 min
Response: 51984
Conc: 1.29 ng/ml



#18 AR-1242-3


R.T.: 4.907 min
Delta R.T.: 0.001 min
Response: 29789
Conc: 2.32 ng/ml



#19 AR-1242-4


R.T.: 5.828 min
Delta R.T.: -0.006 min
Response: 97604
Conc: 3.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.981 min
Delta R.T.: -0.008 min
Response: 122345
Conc: 8.69 ng/ml




#20 AR-1242-5

R.T.: 6.581 min
Delta R.T.: 0.007 min
Response: 740590
Conc: 24.01 ng/ml



#20 AR-1242-5


R.T.: 5.543 min
Delta R.T.: 0.031 min
Response: 131750
Conc: 7.88 ng/ml



#21 AR-1248-1


R.T.: 5.647 min
Delta R.T.: -0.003 min
Response: 22063
Conc: 0.65 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.720 min
Delta R.T.: 0.008 min
Response: 89342
Conc: 6.45 ng/ml




#22 AR-1248-2

R.T.: 5.912 min
Delta R.T.: -0.012 min
Response: 105971
Conc: 2.02 ng/ml



#22 AR-1248-2


R.T.: 4.935 min
Delta R.T.: -0.014 min
Response: 23578
Conc: 1.19 ng/ml



#23 AR-1248-3


R.T.: 6.131 min
Delta R.T.: 0.000 min
Response: 211521
Conc: 3.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.981 min
Delta R.T.: -0.009 min
Response: 122345
Conc: 5.87 ng/ml




#24 AR-1248-4

R.T.: 6.532 min
Delta R.T.: -0.003 min
Response: 647379
Conc: 11.87 ng/ml



#24 AR-1248-4


R.T.: 5.143 min
Delta R.T.: -0.018 min
Response: 87037
Conc: 3.54 ng/ml



#25 AR-1248-5


R.T.: 6.581 min
Delta R.T.: 0.007 min
Response: 740590
Conc: 13.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.561 min
Delta R.T.: 0.007 min
Response: 29930
Conc: 1.38 ng/ml




#26 AR-1254-1

R.T.: 6.532 min
Delta R.T.: 0.023 min
Response: 647379
Conc: 9.83 ng/ml



#26 AR-1254-1


R.T.: 5.543 min
Delta R.T.: 0.031 min
Response: 131750
Conc: 3.72 ng/ml



#27 AR-1254-2


R.T.: 6.744 min
Delta R.T.: 0.016 min
Response: 671665
Conc: 6.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.663 min
Delta R.T.: 0.003 min
Response: 168074
Conc: 5.25 ng/ml




#28 AR-1254-3

R.T.: 7.101 min
Delta R.T.: 0.006 min
Response: 264500
Conc: 2.96 ng/ml



#28 AR-1254-3


R.T.: 6.059 min
Delta R.T.: -0.004 min
Response: 302552
Conc: 6.19 ng/ml



#29 AR-1254-4


R.T.: 7.392 min
Delta R.T.: 0.011 min
Response: 55735
Conc: 1.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.270 min
Delta R.T.: -0.022 min
Response: 50404
Conc: 1.92 ng/ml




#30 AR-1254-5

R.T.: 7.798 min
Delta R.T.: -0.003 min
Response: 263007
Conc: 3.91 ng/ml



#30 AR-1254-5


R.T.: 6.669 min
Delta R.T.: -0.040 min
Response: 33781
Conc: 0.80 ng/ml



#31 AR-1260-1


R.T.: 7.280 min
Delta R.T.: 0.020 min
Response: 66782
Conc: 0.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.172 min
Delta R.T.: -0.023 min
Response: 245955
Conc: 6.97 ng/ml




#32 AR-1260-2

R.T.: 7.516 min
Delta R.T.: -0.001 min
Response: 376696
Conc: 4.77 ng/ml



#32 AR-1260-2


R.T.: 6.371 min
Delta R.T.: -0.012 min
Response: 115906
Conc: 2.89 ng/ml



#33 AR-1260-3


R.T.: 7.868 min
Delta R.T.: -0.009 min
Response: 152950
Conc: 2.80 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.545 min
Delta R.T.: 0.008 min
Response: 33193
Conc: 0.88 ng/ml




#34 AR-1260-4

R.T.: 8.106 min
Delta R.T.: 0.001 min
Response: 382698
Conc: 6.11 ng/ml



#34 AR-1260-4


R.T.: 7.018 min
Delta R.T.: 0.010 min
Response: 186070
Conc: 6.44 ng/ml



#35 AR-1260-5


R.T.: 8.430 min
Delta R.T.: 0.002 min
Response: 339452
Conc: 3.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.250 min
Delta R.T.: 0.000 min
Response: 28212
Conc: 0.47 ng/ml




#36 AR-1262-1

R.T.: 7.868 min
Delta R.T.: -0.008 min
Response: 152950
Conc: 1.69 ng/ml



#36 AR-1262-1


R.T.: 6.780 min
Delta R.T.: 0.034 min
Response: 396847
Conc: 8.46 ng/ml



#37 AR-1262-2


R.T.: 8.430 min
Delta R.T.: 0.003 min
Response: 339452
Conc: 2.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.018 min
Delta R.T.: 0.011 min
Response: 186070
Conc: 4.48 ng/ml




#38 AR-1262-3

R.T.: 8.771 min
Delta R.T.: 0.026 min
Response: -217954
Conc: N.D.



#38 AR-1262-3


R.T.: 7.522 min
Delta R.T.: -0.009 min
Response: 90729
Conc: 2.84 ng/ml



#39 AR-1262-4


R.T.: 8.825 min
Delta R.T.: -0.011 min
Response: 210167
Conc: 2.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.617 min
Delta R.T.: 0.022 min
Response: 111591
Conc: 2.03 ng/ml




#40 AR-1262-5

R.T.: 9.509 min
Delta R.T.: 0.003 min
Response: 36433
Conc: 0.83 ng/ml



#40 AR-1262-5


R.T.: 8.049 min
Delta R.T.: -0.041 min
Response: 27686
Conc: 1.10 ng/ml



#41 AR-1268-1


R.T.: 8.771 min
Delta R.T.: 0.027 min
Response: -217954
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.522 min
Delta R.T.: -0.010 min
Response: 90729
Conc: 1.02 ng/ml




#42 AR-1268-2

R.T.: 8.856 min
Delta R.T.: 0.015 min
Response: 264037
Conc: 1.68 ng/ml



#42 AR-1268-2


R.T.: 7.617 min
Delta R.T.: 0.021 min
Response: 111591
Conc: 1.41 ng/ml



#43 AR-1268-3


R.T.: 9.074 min
Delta R.T.: 0.002 min
Response: 133800
Conc: 0.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.840 min
Delta R.T.: 0.038 min
Response: 125427
Conc: 1.78 ng/ml




#44 AR-1268-4

R.T.: 9.509 min
Delta R.T.: 0.002 min
Response: 36433
Conc: 0.73 ng/ml



#44 AR-1268-4


R.T.: 8.049 min
Delta R.T.: -0.042 min
Response: 27686
Conc: 1.00 ng/ml



#45 AR-1268-5

R.T.: 9.929 min
Delta R.T.: -0.004 min
Response: 364755
Conc: 0.89 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.372 min
Delta R.T.: -0.006 min
Response: 183894
Conc: 0.96 ng/ml