

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103023\
 Data File : PO099122.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2023 10:00
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 11:24:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.476	3.627	87140203	32456997	44.101	45.099
2) SA Decachlor...	10.281	8.630	58074911	24276061	51.264	45.246
Target Compounds						
21) L5 AR-1248-1	5.650	4.708	16088160	6667494	473.562	481.061
22) L5 AR-1248-2	5.925	4.945	24022831	9108822	459.035	458.569
23) L5 AR-1248-3	6.130	4.986	25587076	9709474	449.072	465.603
24) L5 AR-1248-4	6.534	5.158	25824536	12141382	473.570	493.883
25) L5 AR-1248-5	6.574	5.551	26205075	11062184	461.307	509.346



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103023\
Data File : PO099122.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2023 10:00
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 11:24:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.476 min
Delta R.T.: -0.002 min
Response: 87140203
Conc: 44.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.627 min
Delta R.T.: -0.004 min
Response: 32456997
Conc: 45.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.281 min
Delta R.T.: 0.000 min
Response: 58074911
Conc: 51.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.630 min
Delta R.T.: 0.000 min
Response: 24276061
Conc: 45.25 ng/ml



#21 AR-1248-1


R.T.: 5.650 min
Delta R.T.: 0.000 min
Response: 16088160
Conc: 473.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.708 min
Delta R.T.: -0.003 min
Response: 6667494
Conc: 481.06 ng/ml




#22 AR-1248-2

R.T.: 5.925 min
Delta R.T.: 0.000 min
Response: 24022831
Conc: 459.04 ng/ml



#22 AR-1248-2


R.T.: 4.945 min
Delta R.T.: -0.003 min
Response: 9108822
Conc: 458.57 ng/ml



#23 AR-1248-3


R.T.: 6.130 min
Delta R.T.: 0.000 min
Response: 25587076
Conc: 449.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.986 min
Delta R.T.: -0.004 min
Response: 9709474
Conc: 465.60 ng/ml




#24 AR-1248-4

R.T.: 6.534 min
Delta R.T.: 0.000 min
Response: 25824536
Conc: 473.57 ng/ml



#24 AR-1248-4


R.T.: 5.158 min
Delta R.T.: -0.003 min
Response: 12141382
Conc: 493.88 ng/ml



#25 AR-1248-5

R.T.: 6.574 min
Delta R.T.: 0.000 min
Response: 26205075
Conc: 461.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.551 min
Delta R.T.: -0.003 min
Response: 11062184
Conc: 509.35 ng/ml