

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0103123\
 Data File : P0099217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2023 21:47
 Operator : YP/AJ
 Sample : 05161-01 10X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC147245-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/01/2023
 Supervised By :mohammad ahmed 11/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 00:51:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 25 06:04:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.471	3.626	3736083	1233039	1.891	1.713
2) SA Decachlor...	10.270	8.622	2707889	1030871	2.390m	1.921

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0103123\
Data File : P0099217.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2023 21:47
Operator : YP/AJ
Sample : 05161-01 10X
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
BC147245-1-1


Manual Integrations

Reviewed By :Yogesh Patel 11/01/2023
Supervised By :mohammad ahmed 11/01/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 00:51:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102523.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 25 06:04:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.471 min
Delta R.T.: -0.007 min
Response: 3736083
Conc: 1.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC147245-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/01/2023
Supervised By :mohammad ahmed 11/01/2023




#1 Tetrachloro-m-xylene

R.T.: 3.626 min
Delta R.T.: -0.006 min
Response: 1233039
Conc: 1.71 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.270 min
Delta R.T.: -0.010 min
Response: 2707889
Conc: 2.39 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.622 min
Delta R.T.: -0.008 min
Response: 1030871
Conc: 1.92 ng/ml