

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103124\
 Data File : PO107575.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2024 20:09
 Operator : YP/AJ
 Sample : P4620-30
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P001-CF16-02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 23:15:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Oct 28 11:34:55 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.371	3.644	194.1E6	69883512	21.302	21.713
2)	SA Decachlor...	10.069	8.639	45478668	50521925	18.537	18.428
Target Compounds							
3)	L1 AR-1016-1	5.527	4.728	407886	339336	1.512	3.286 #
4)	L1 AR-1016-2	5.545	4.760	2066889	77177	5.208	0.547 #
5)	L1 AR-1016-3	5.596	4.921	3235758	255653	12.883	3.247 #
6)	L1 AR-1016-4	5.707	4.962	499222	565703	2.592	8.557 #
7)	L1 AR-1016-5	6.010	5.166	795985	108196	4.358	1.317 #
8)	L2 AR-1221-1	4.564	3.847	2301565	109739	20.638	3.028 #
9)	L2 AR-1221-2	4.671	3.945	4388682	1602328	53.575	55.056
10)	L2 AR-1221-3	4.734	4.019	137708	293677	0.575	3.328 #
11)	L3 AR-1232-1	4.734	4.019	137708	293677	0.703	4.362 #
12)	L3 AR-1232-2	5.260	4.760	646997	77177	6.280	1.221 #
13)	L3 AR-1232-3	5.545	4.921	2066889	255653	11.547	7.416 #
14)	L3 AR-1232-4	5.707	5.001	499222	122581	5.537	3.896 #
15)	L3 AR-1232-5	5.793	5.166	696390	108196	10.567	3.270 #
16)	L4 AR-1242-1	5.527	4.728	407886	339336	1.824	4.052 #
17)	L4 AR-1242-2	5.545	4.760	2066889	77177	6.389	0.667 #
18)	L4 AR-1242-3	5.596	4.921	3235758	255653	15.484	4.030 #
19)	L4 AR-1242-4	5.707	5.001	499222	122581	3.120	1.914 #
20)	L4 AR-1242-5	6.441	5.534	708474	154840	4.528	1.997 #
21)	L5 AR-1248-1	5.527	4.728	407886	339336	2.346	5.349 #
22)	L5 AR-1248-2	5.793	4.962	696390	565703	2.752	6.184 #
23)	L5 AR-1248-3	6.010	5.001	795985	122581	3.113	1.285 #
24)	L5 AR-1248-4	6.397	5.166	959338	108196	3.721	0.965 #
25)	L5 AR-1248-5	6.441	5.559	708474	214774	2.724	1.948 #
26)	L6 AR-1254-1	6.384	5.534	527872	154840	1.964	0.940 #
27)	L6 AR-1254-2	6.593	5.623f	1237939	169869	3.280	1.172 #
28)	L6 AR-1254-3	6.945	6.060	2070119	203317	5.786	0.885 #
29)	L6 AR-1254-4	7.249	6.302	464926	142079	2.027	1.098 #
30)	L6 AR-1254-5	7.656	6.714	390951	737656	1.883	3.861 #
31)	L7 AR-1260-1	7.110	6.207	621914	180434	2.405	1.163 #
32)	L7 AR-1260-2	7.380	6.379	358390	640496	1.362	3.637 #
33)	L7 AR-1260-3	7.745	6.553	674675	189174	3.754	1.128 #
34)	L7 AR-1260-4	7.963	7.010	846537	147257	4.823	1.018 #
35)	L7 AR-1260-5	8.258	7.249	1133511	1379502	3.986	4.187
36)	L8 AR-1262-1	7.745	6.755	674675	185532	2.748	0.872 #
37)	L8 AR-1262-2	8.258	7.249	1133511	1379502	3.663	3.797
38)	L8 AR-1262-3	8.589	7.531	53695	275364	0.257	2.004 #
39)	L8 AR-1262-4	8.678	7.609	431662	32236	2.647	0.122 #
40)	L8 AR-1262-5	9.318	8.101	31641	421933	0.308	3.739 #
41)	L9 AR-1268-1	8.589	7.531	53695	275364	0.148	0.681 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103124\
Data File : PO107575.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2024 20:09
Operator : YP/AJ
Sample : P4620-30
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 23:15:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101524.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 28 11:34:55 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.678	7.609	431662	32236	1.339	0.086 #
43)	L9 AR-1268-3	8.914	7.803	124808	37923	0.447	0.114 #
44)	L9 AR-1268-4	9.318	8.101	31641	421933	0.284	3.366 #
45)	L9 AR-1268-5	9.736	8.385	461497	384296	0.554	0.389 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103124\
Data File : PO107575.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2024 20:09
Operator : YP/AJ
Sample : P4620-30
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 23:15:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101524.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Oct 28 11:34:55 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.371 min
Delta R.T.: -0.003 min
Response: 194147361
Conc: 21.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: 0.000 min
Response: 69883512
Conc: 21.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.069 min
Delta R.T.: 0.011 min
Response: 45478668
Conc: 18.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.639 min
Delta R.T.: 0.000 min
Response: 50521925
Conc: 18.43 ng/ml



#3 AR-1016-1


R.T.: 5.527 min
Delta R.T.: 0.005 min
Response: 407886
Conc: 1.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#3 AR-1016-1

R.T.: 4.728 min
Delta R.T.: 0.002 min
Response: 339336
Conc: 3.29 ng/ml




#4 AR-1016-2

R.T.: 5.545 min
Delta R.T.: 0.000 min
Response: 2066889
Conc: 5.21 ng/ml



#4 AR-1016-2


R.T.: 4.760 min
Delta R.T.: 0.014 min
Response: 77177
Conc: 0.55 ng/ml



#5 AR-1016-3


R.T.: 5.596 min
Delta R.T.: -0.012 min
Response: 3235758
Conc: 12.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#5 AR-1016-3

R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 255653
Conc: 3.25 ng/ml




#6 AR-1016-4

R.T.: 5.707 min
Delta R.T.: 0.003 min
Response: 499222
Conc: 2.59 ng/ml



#6 AR-1016-4


R.T.: 4.962 min
Delta R.T.: 0.000 min
Response: 565703
Conc: 8.56 ng/ml



#7 AR-1016-5


R.T.: 6.010 min
Delta R.T.: 0.011 min
Response: 795985
Conc: 4.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#7 AR-1016-5

R.T.: 5.166 min
Delta R.T.: -0.010 min
Response: 108196
Conc: 1.32 ng/ml




#8 AR-1221-1

R.T.: 4.564 min
Delta R.T.: -0.007 min
Response: 2301565
Conc: 20.64 ng/ml



#8 AR-1221-1


R.T.: 3.847 min
Delta R.T.: -0.009 min
Response: 109739
Conc: 3.03 ng/ml



#9 AR-1221-2


R.T.: 4.671 min
Delta R.T.: 0.015 min
Response: 4388682
Conc: 53.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#9 AR-1221-2

R.T.: 3.945 min
Delta R.T.: 0.004 min
Response: 1602328
Conc: 55.06 ng/ml




#10 AR-1221-3

R.T.: 4.734 min
Delta R.T.: 0.003 min
Response: 137708
Conc: 0.58 ng/ml



#10 AR-1221-3


R.T.: 4.019 min
Delta R.T.: 0.002 min
Response: 293677
Conc: 3.33 ng/ml



#11 AR-1232-1


R.T.: 4.734 min
Delta R.T.: 0.000 min
Response: 137708
Conc: 0.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#11 AR-1232-1

R.T.: 4.019 min
Delta R.T.: 0.001 min
Response: 293677
Conc: 4.36 ng/ml




#12 AR-1232-2

R.T.: 5.260 min
Delta R.T.: 0.002 min
Response: 646997
Conc: 6.28 ng/ml



#12 AR-1232-2


R.T.: 4.760 min
Delta R.T.: 0.014 min
Response: 77177
Conc: 1.22 ng/ml



#13 AR-1232-3


R.T.: 5.545 min
Delta R.T.: 0.000 min
Response: 2066889
Conc: 11.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#13 AR-1232-3

R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 255653
Conc: 7.42 ng/ml




#14 AR-1232-4

R.T.: 5.707 min
Delta R.T.: 0.003 min
Response: 499222
Conc: 5.54 ng/ml



#14 AR-1232-4


R.T.: 5.001 min
Delta R.T.: -0.004 min
Response: 122581
Conc: 3.90 ng/ml



#15 AR-1232-5


R.T.: 5.793 min
Delta R.T.: -0.003 min
Response: 696390
Conc: 10.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#15 AR-1232-5

R.T.: 5.166 min
Delta R.T.: -0.011 min
Response: 108196
Conc: 3.27 ng/ml




#16 AR-1242-1

R.T.: 5.527 min
Delta R.T.: 0.005 min
Response: 407886
Conc: 1.82 ng/ml



#16 AR-1242-1


R.T.: 4.728 min
Delta R.T.: 0.002 min
Response: 339336
Conc: 4.05 ng/ml



#17 AR-1242-2


R.T.: 5.545 min
Delta R.T.: 0.000 min
Response: 2066889
Conc: 6.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#17 AR-1242-2

R.T.: 4.760 min
Delta R.T.: 0.014 min
Response: 77177
Conc: 0.67 ng/ml




#18 AR-1242-3

R.T.: 5.596 min
Delta R.T.: -0.012 min
Response: 3235758
Conc: 15.48 ng/ml



#18 AR-1242-3


R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 255653
Conc: 4.03 ng/ml



#19 AR-1242-4


R.T.: 5.707 min
Delta R.T.: 0.003 min
Response: 499222
Conc: 3.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#19 AR-1242-4

R.T.: 5.001 min
Delta R.T.: -0.003 min
Response: 122581
Conc: 1.91 ng/ml




#20 AR-1242-5

R.T.: 6.441 min
Delta R.T.: 0.004 min
Response: 708474
Conc: 4.53 ng/ml



#20 AR-1242-5


R.T.: 5.534 min
Delta R.T.: 0.007 min
Response: 154840
Conc: 2.00 ng/ml



#21 AR-1248-1


R.T.: 5.527 min
Delta R.T.: 0.006 min
Response: 407886
Conc: 2.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#21 AR-1248-1

R.T.: 4.728 min
Delta R.T.: 0.002 min
Response: 339336
Conc: 5.35 ng/ml




#22 AR-1248-2

R.T.: 5.793 min
Delta R.T.: -0.003 min
Response: 696390
Conc: 2.75 ng/ml



#22 AR-1248-2


R.T.: 4.962 min
Delta R.T.: 0.000 min
Response: 565703
Conc: 6.18 ng/ml



#23 AR-1248-3


R.T.: 6.010 min
Delta R.T.: 0.012 min
Response: 795985
Conc: 3.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#23 AR-1248-3

R.T.: 5.001 min
Delta R.T.: -0.003 min
Response: 122581
Conc: 1.29 ng/ml




#24 AR-1248-4

R.T.: 6.397 min
Delta R.T.: 0.000 min
Response: 959338
Conc: 3.72 ng/ml



#24 AR-1248-4


R.T.: 5.166 min
Delta R.T.: -0.010 min
Response: 108196
Conc: 0.96 ng/ml



#25 AR-1248-5


R.T.: 6.441 min
Delta R.T.: 0.004 min
Response: 708474
Conc: 2.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#25 AR-1248-5

R.T.: 5.559 min
Delta R.T.: -0.008 min
Response: 214774
Conc: 1.95 ng/ml




#26 AR-1254-1

R.T.: 6.384 min
Delta R.T.: 0.010 min
Response: 527872
Conc: 1.96 ng/ml



#26 AR-1254-1


R.T.: 5.534 min
Delta R.T.: 0.007 min
Response: 154840
Conc: 0.94 ng/ml



#27 AR-1254-2


R.T.: 6.593 min
Delta R.T.: 0.000 min
Response: 1237939
Conc: 3.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#27 AR-1254-2

R.T.: 5.623 min
Delta R.T.: -0.051 min
Response: 169869
Conc: 1.17 ng/ml




#28 AR-1254-3

R.T.: 6.945 min
Delta R.T.: -0.012 min
Response: 2070119
Conc: 5.79 ng/ml



#28 AR-1254-3


R.T.: 6.060 min
Delta R.T.: -0.017 min
Response: 203317
Conc: 0.89 ng/ml



#29 AR-1254-4


R.T.: 7.249 min
Delta R.T.: 0.007 min
Response: 464926
Conc: 2.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#29 AR-1254-4

R.T.: 6.302 min
Delta R.T.: -0.003 min
Response: 142079
Conc: 1.10 ng/ml




#30 AR-1254-5

R.T.: 7.656 min
Delta R.T.: -0.007 min
Response: 390951
Conc: 1.88 ng/ml



#30 AR-1254-5


R.T.: 6.714 min
Delta R.T.: -0.007 min
Response: 737656
Conc: 3.86 ng/ml



#31 AR-1260-1


R.T.: 7.110 min
Delta R.T.: -0.015 min
Response: 621914
Conc: 2.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#31 AR-1260-1

R.T.: 6.207 min
Delta R.T.: 0.000 min
Response: 180434
Conc: 1.16 ng/ml




#32 AR-1260-2

R.T.: 7.380 min
Delta R.T.: -0.002 min
Response: 358390
Conc: 1.36 ng/ml



#32 AR-1260-2


R.T.: 6.379 min
Delta R.T.: -0.015 min
Response: 640496
Conc: 3.64 ng/ml



#33 AR-1260-3


R.T.: 7.745 min
Delta R.T.: 0.002 min
Response: 674675
Conc: 3.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#33 AR-1260-3

R.T.: 6.553 min
Delta R.T.: 0.005 min
Response: 189174
Conc: 1.13 ng/ml




#34 AR-1260-4

R.T.: 7.963 min
Delta R.T.: -0.005 min
Response: 846537
Conc: 4.82 ng/ml



#34 AR-1260-4


R.T.: 7.010 min
Delta R.T.: -0.008 min
Response: 147257
Conc: 1.02 ng/ml



#35 AR-1260-5


R.T.: 8.258 min
Delta R.T.: -0.023 min
Response: 1133511
Conc: 3.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#35 AR-1260-5

R.T.: 7.249 min
Delta R.T.: -0.010 min
Response: 1379502
Conc: 4.19 ng/ml




#36 AR-1262-1

R.T.: 7.745 min
Delta R.T.: 0.003 min
Response: 674675
Conc: 2.75 ng/ml



#36 AR-1262-1


R.T.: 6.755 min
Delta R.T.: -0.004 min
Response: 185532
Conc: 0.87 ng/ml



#37 AR-1262-2


R.T.: 8.258 min
Delta R.T.: -0.021 min
Response: 1133511
Conc: 3.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#37 AR-1262-2

R.T.: 7.249 min
Delta R.T.: -0.009 min
Response: 1379502
Conc: 3.80 ng/ml




#38 AR-1262-3

R.T.: 8.589 min
Delta R.T.: -0.003 min
Response: 53695
Conc: 0.26 ng/ml



#38 AR-1262-3


R.T.: 7.531 min
Delta R.T.: -0.010 min
Response: 275364
Conc: 2.00 ng/ml



#39 AR-1262-4


R.T.: 8.678 min
Delta R.T.: 0.000 min
Response: 431662
Conc: 2.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#39 AR-1262-4

R.T.: 7.609 min
Delta R.T.: 0.004 min
Response: 32236
Conc: 0.12 ng/ml




#40 AR-1262-5

R.T.: 9.318 min
Delta R.T.: 0.003 min
Response: 31641
Conc: 0.31 ng/ml



#40 AR-1262-5


R.T.: 8.101 min
Delta R.T.: 0.004 min
Response: 421933
Conc: 3.74 ng/ml



#41 AR-1268-1


R.T.: 8.589 min
Delta R.T.: 0.002 min
Response: 53695
Conc: 0.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#41 AR-1268-1

R.T.: 7.531 min
Delta R.T.: -0.010 min
Response: 275364
Conc: 0.68 ng/ml




#42 AR-1268-2

R.T.: 8.678 min
Delta R.T.: -0.001 min
Response: 431662
Conc: 1.34 ng/ml



#42 AR-1268-2


R.T.: 7.609 min
Delta R.T.: 0.003 min
Response: 32236
Conc: 0.09 ng/ml



#43 AR-1268-3


R.T.: 8.914 min
Delta R.T.: 0.006 min
Response: 124808
Conc: 0.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02




#43 AR-1268-3

R.T.: 7.803 min
Delta R.T.: -0.008 min
Response: 37923
Conc: 0.11 ng/ml




#44 AR-1268-4

R.T.: 9.318 min
Delta R.T.: 0.004 min
Response: 31641
Conc: 0.28 ng/ml



#44 AR-1268-4


R.T.: 8.101 min
Delta R.T.: 0.003 min
Response: 421933
Conc: 3.37 ng/ml



#45 AR-1268-5

R.T.: 9.736 min
Delta R.T.: 0.014 min
Response: 461497
Conc: 0.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
P001-CF16-02



#45 AR-1268-5

R.T.: 8.385 min
Delta R.T.: 0.000 min
Response: 384296
Conc: 0.39 ng/ml