

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0103125\
 Data File : P0114830.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2025 22:24
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/03/2025
 Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 23:22:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0100925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 02:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.350	3.652	291.7E6	194.7E6	53.728	48.647
2) SA Decachlor...	9.939	8.649	182.7E6	170.1E6	50.176	51.039
Target Compounds						
21) L5 AR-1248-1	5.489	4.731	54000105	44745960	549.014	454.295m
22) L5 AR-1248-2	5.760	4.967	75738487	58794705	527.556	438.590
23) L5 AR-1248-3	5.960	5.009	82625518	62777776	520.274	449.885
24) L5 AR-1248-4	6.357	5.179	96618260	74717638	493.478	454.586
25) L5 AR-1248-5	6.395	5.571	91479665	75949688	456.955	447.874

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO103125\
Data File : PO114830.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 22:24
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 23:22:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO100925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 02:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.350 min
Delta R.T.: -0.005 min
Response: 291685749
Conc: 53.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#1 Tetrachloro-m-xylene

R.T.: 3.652 min
Delta R.T.: -0.007 min
Response: 194699630
Conc: 48.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.939 min
Delta R.T.: 0.005 min
Response: 182698213
Conc: 50.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.649 min
Delta R.T.: -0.023 min
Response: 170122202
Conc: 51.04 ng/ml




#21 AR-1248-1

R.T.: 5.489 min
Delta R.T.: -0.004 min
Response: 54000105
Conc: 549.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#21 AR-1248-1

R.T.: 4.731 min
Delta R.T.: -0.011 min
Response: 44745960
Conc: 454.29 ng/ml




#22 AR-1248-2

R.T.: 5.760 min
Delta R.T.: -0.002 min
Response: 75738487
Conc: 527.56 ng/ml



#22 AR-1248-2

R.T.: 4.967 min
Delta R.T.: -0.012 min
Response: 58794705
Conc: 438.59 ng/ml




#23 AR-1248-3

R.T.: 5.960 min
Delta R.T.: -0.003 min
Response: 82625518
Conc: 520.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#23 AR-1248-3

R.T.: 5.009 min
Delta R.T.: -0.012 min
Response: 62777776
Conc: 449.89 ng/ml




#24 AR-1248-4

R.T.: 6.357 min
Delta R.T.: -0.002 min
Response: 96618260
Conc: 493.48 ng/ml



#24 AR-1248-4

R.T.: 5.179 min
Delta R.T.: -0.012 min
Response: 74717638
Conc: 454.59 ng/ml




#25 AR-1248-5

R.T.: 6.395 min
Delta R.T.: -0.002 min
Response: 91479665
Conc: 456.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025



#25 AR-1248-5

R.T.: 5.571 min
Delta R.T.: -0.012 min
Response: 75949688
Conc: 447.87 ng/ml