

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110118\
 Data File : P0050862.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Nov 2018 8:45
 Operator : SM/SJ
 Sample : J5719-01RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S1RE

Manual Integrations
 APPROVED

Sohil
 11/2/2018 6:23:36 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 14:23:20 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 30 07:48:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.352	3.347	3061847	1416807	6.125m	6.518
2) SA Decachlor...	10.008	8.090	2847688	2182364	9.162	8.532

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110118\
Data File : PO050862.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Nov 2018 8:45
Operator : SM/SJ
Sample : J5719-01RE
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
S1RE


Manual Integrations
APPROVED

Sohil
11/2/2018 6:23:36 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 14:23:20 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102918.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 30 07:48:38 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.352 min
Delta R.T.: -0.004 min
Response: 3061847
Conc: 6.13 ng/ml m

Instrument :
ECD_O
ClientSampleId :
S1RE


Manual Integrations
APPROVED

Sohil
11/2/2018 6:23:36 PM




#1 Tetrachloro-m-xylene

R.T.: 3.347 min
Delta R.T.: 0.000 min
Response: 1416807
Conc: 6.52 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.008 min
Delta R.T.: -0.009 min
Response: 2847688
Conc: 9.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.090 min
Delta R.T.: -0.005 min
Response: 2182364
Conc: 8.53 ng/ml